

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060419\
 Data File : VV011218.D
 Acq On : 04 Jun 2019 20:55
 Operator : SY/MD
 Sample : K3081-19
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GALE7

Quant Time: Jun 05 02:36:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jun 05 01:41:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	174463	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	161742	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	67496	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	61920	4.68	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.60%
7) Chloroethane-d5	1.56	69	48423	4.41	ug/L	-0.02
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.20%
11) 1,1-Dichloroethene-d2	2.12	63	95198	3.60	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.00%
20) 2-Butanone-d5	3.94	46	162767	54.69	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.38%
24) Chloroform-d	4.40	84	108704	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.40%
26) 1,2-Dichloroethane-d4	5.08	65	62749	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.80%
32) Benzene-d6	5.09	84	227712	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.20%
36) 1,2-Dichloropropane-d6	6.12	67	69920	4.69	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.80%
41) Toluene-d8	7.36	98	205412	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.40%
43) trans-1,3-Dichloropropene-	7.66	79	24431	4.30	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.00%
46) 2-Hexanone-d5	8.13	63	133366	55.56	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	111.12%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	50934	5.01	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.20%
64) 1,2-Dichlorobenzene-d4	11.67	152	67139	5.04	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.80%

Target Compounds					Ovalue
13) Acetone	2.19	43	23588	11.821 ug/L	95
21) 2-Butanone	4.04	43	8140	2.649 ug/L	98
42) Toluene	7.43	91	11587	0.211 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV060419\
Data File : VV011218.D
Acq On : 04 Jun 2019 20:55
Operator : SY/MD
Sample : K3081-19
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GALE7

Quant Time: Jun 05 02:36:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jun 05 01:41:25 2019
Response via : Initial Calibration

