

Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV060419\
 Data File : VV011207.D
 Acq On : 04 Jun 2019 15:25
 Operator : SY/MD
 Sample : VSTDICV005
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV42

Quant Time: Jun 05 01:37:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jun 04 15:14:23 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	1.000	1.000	0.0	116	0.00
2 T	Dichlorodifluoromethane	0.493	0.415	15.8	101	0.00
3 T	Chloromethane	0.477	0.433	9.2	109	0.00
4 S	Vinyl Chloride-d3	0.379	0.318	16.1	100	0.00
5 T	Vinyl chloride	0.460	0.416	9.6	107	0.00
6 T	Bromomethane	0.239	0.221	7.5	110	0.00
7 S	Chloroethane-d5	0.315	0.250	20.6	98	0.00
8 T	Chloroethane	0.294	0.271	7.8	111	0.00
9 T	Trichlorofluoromethane	0.631	0.577	8.6	110	0.00
10 T	1,1,2-Trichloro-1,2,2-trifl	0.324	0.298	8.0	109	0.00
11 S	1,1-Dichloroethene-d2	0.757	0.657	13.2	102	0.00
12 T	1,1-Dichloroethene	0.315	0.284	9.8	112	0.00
13 T	Acetone	0.057	0.067	-17.5	138	0.00
14 T	Carbon disulfide	0.956	0.878	8.2	111	0.00
15 T	Methyl Acetate	0.152	0.139	8.6	114	0.00
16 T	Methylene chloride	0.325	0.313	3.7	117	0.00
17 T	Methyl tert-butyl Ether	0.841	0.775	7.8	109	0.00
18 T	trans-1,2-Dichloroethene	0.353	0.333	5.7	111	0.00
19 T	1,1-Dichloroethane	0.696	0.652	6.3	113	0.00
20 S	2-Butanone-d5	0.085	0.073	14.1	97	0.00
21 T	2-Butanone	0.088	0.096	-9.1	127	0.00
22 T	cis-1,2-Dichloroethene	0.391	0.369	5.6	115	0.00
23 T	Bromochloromethane	0.150	0.142	5.3	110	0.00
24 S	Chloroform-d	0.667	0.570	14.5	100	0.00
25 T	Chloroform	0.719	0.668	7.1	111	0.00
26 S	1,2-Dichloroethane-d4	0.360	0.301	16.4	97	0.00
27 T	1,2-Dichloroethane	0.439	0.417	5.0	115	0.00
28 I	Chlorobenzene-d5	1.000	1.000	0.0	113	0.00
29 T	1,1,1-Trichloroethane	0.638	0.616	3.4	114	0.00
30 T	Cyclohexane	0.707	0.667	5.7	112	0.00
31 T	Carbon tetrachloride	0.533	0.516	3.2	115	0.00
32 S	Benzene-d6	1.496	1.296	13.4	100	0.00
33 T	Benzene	1.591	1.549	2.6	111	0.00
34 T	Trichloroethene	0.411	0.409	0.5	116	0.00
35 T	Methylcyclohexane	0.680	0.644	5.3	113	0.00
36 S	1,2-Dichloropropane-d6	0.461	0.391	15.2	98	0.00
37 T	1,2-Dichloropropane	0.396	0.387	2.3	113	0.00
38 T	Bromodichloromethane	0.490	0.487	0.6	116	0.00
39 T	cis-1,3-Dichloropropene	0.579	0.574	0.9	114	0.00
40 T	4-Methyl-2-pentanone	0.252	0.258	-2.4	117	0.00
41 S	Toluene-d8	1.374	1.216	11.5	101	0.00
42 T	Toluene	1.696	1.666	1.8	112	0.00
43 S	trans-1,3-Dichloropropene-d	0.176	0.150	14.8	98	0.00
44 T	trans-1,3-Dichloropropene	0.440	0.446	-1.4	118	0.00
45 T	1,1,2-Trichloroethane	0.253	0.253	0.0	117	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV060419\
 Data File : VV011207.D
 Acq On : 04 Jun 2019 15:25
 Operator : SY/MD
 Sample : VSTDICV005
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV42

Quant Time: Jun 05 01:37:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jun 04 15:14:23 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 S	2-Hexanone-d5	0.074	0.065	12.2	99	0.00
47 T	Tetrachloroethene	0.297	0.301	-1.3	119	0.00
48 T	2-Hexanone	0.176	0.186	-5.7	119	0.00
49 T	Dibromochloromethane	0.288	0.284	1.4	111	0.00
50 T	1,2-Dibromoethane	0.241	0.228	5.4	108	0.00
51 T	Chlorobenzene	1.034	1.013	2.0	113	0.00
52 T	Ethylbenzene	1.883	1.887	-0.2	115	0.00
53 T	m,p-xylene	0.696	0.698	-0.3	116	0.00
54 T	o-xylene	0.683	0.690	-1.0	115	0.00
55 T	Styrene	1.113	1.141	-2.5	113	0.00
56 T	Isopropylbenzene	1.839	1.832	0.4	116	0.00
57 S	1,1,2,2-Tetrachloroethane-d	0.314	0.267	15.0	99	0.00
58 T	1,1,2,2-Tetrachloroethane	0.307	0.292	4.9	109	0.00
59	1,2,3-Trichloropropane	0.231	0.228	1.3	115	0.00
60 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	113	0.00
61 T	Bromoform	0.300	0.304	-1.3	118	0.00
62 T	1,3-Dichlorobenzene	1.653	1.619	2.1	115	0.00
63 T	1,4-Dichlorobenzene	1.605	1.627	-1.4	121	0.00
64 S	1,2-Dichlorobenzene-d4	0.987	0.847	14.2	98	0.00
65 T	1,2-Dichlorobenzene	1.528	1.536	-0.5	117	0.00
66 T	1,2-Dibromo-3-chloropropane	0.110	0.100	9.1	111	0.00
67	1,3,5-Trichlorobenzene	1.189	1.207	-1.5	119	0.00
68 T	1,2,4-trichlorobenzene	0.903	0.924	-2.3	119	0.00
69	Naphthalene	1.478	1.597	-8.1	123	0.00
70 T	1,2,3-Trichlorobenzene	0.810	0.858	-5.9	119	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0