

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060419\
 Data File : VV011213.D
 Acq On : 04 Jun 2019 18:47
 Operator : SY/MD
 Sample : K2888-15
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BP7

Quant Time: Jun 05 02:14:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jun 05 01:41:25 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	186996	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	170148	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	72626	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	62213	4.39	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.80%
7) Chloroethane-d5	1.56	69	48605	4.13	ug/L	-0.02
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.60%
11) 1,1-Dichloroethene-d2	2.12	63	93417	3.30	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.00%
20) 2-Butanone-d5	3.94	46	163757	51.33	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.66%
24) Chloroform-d	4.40	84	106606	4.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.60%
26) 1,2-Dichloroethane-d4	5.08	65	61237	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.80%
32) Benzene-d6	5.09	84	224189	4.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.00%
36) 1,2-Dichloropropane-d6	6.12	67	69202	4.41	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.20%
41) Toluene-d8	7.36	98	203452	4.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.00%
43) trans-1,3-Dichloropropene-	7.66	79	25124	4.20	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.00%
46) 2-Hexanone-d5	8.13	63	134256	53.17	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.34%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	49305	4.61	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.20%
64) 1,2-Dichlorobenzene-d4	11.67	152	66822	4.66	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.20%

Target Compounds					Ovalue
13) Acetone	2.19	43	9839	4.600 ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#13

Acetone

Concen: 4.600 ug/L

RT: 2.19 min Scan# 342

Delta R.T. -0.03 min

Lab File: VV011213.D

Acq: 04 Jun 2019 18:47

Instrument :
MSVOA_V
ClientSampleId :
C0BP7

Tot Ion: 43 Resp: 9839

Ion Ratio Lower Upper

43 100

58 31.0 0.0 55.2

