

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060519\
 Data File : VV011228.D
 Acq On : 05 Jun 2019 14:37
 Operator : SY/MD
 Sample : K3081-21
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Jun 06 06:36:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jun 06 05:43:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	177155	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	164321	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	68980	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	57686	4.30	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.00%
7) Chloroethane-d5	1.56	69	45985	4.12	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.40%
11) 1,1-Dichloroethene-d2	2.12	63	89161	3.32	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.40%
20) 2-Butanone-d5	3.94	46	138584	45.86	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.72%
24) Chloroform-d	4.40	84	101496	4.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.00%
26) 1,2-Dichloroethane-d4	5.08	65	58187	4.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.20%
32) Benzene-d6	5.09	84	213720	4.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.00%
36) 1,2-Dichloropropane-d6	6.12	67	64521	4.26	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.20%
41) Toluene-d8	7.36	98	197288	4.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.40%
43) trans-1,3-Dichloropropene-	7.66	79	22384	3.88	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.60%
46) 2-Hexanone-d5	8.13	63	109384	44.86	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.72%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	42475	4.11	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.20%
64) 1,2-Dichlorobenzene-d4	11.67	152	60218	4.42	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV060519\
Data File : VV011228.D
Acq On : 05 Jun 2019 14:37
Operator : SY/MD
Sample : K3081-21
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jun 06 06:36:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jun 06 05:43:55 2019
Response via : Initial Calibration

