

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061124\
 Data File : VW036036.D
 Acq On : 11 Jun 2024 15:56
 Operator : SY/MD
 Sample : P2798-06
 Misc : 5.31g/10.0mL/MSVOA_V/SOIL/A
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BHB17

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/12/2024
 Supervised By :Semsettin Yesilyurt 06/12/2024

Quant Time: Jun 12 02:08:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM053024SMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jun 12 02:05:45 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	352037	25.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	375679	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	180248	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	146971	22.708	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery = 90.840%			
7) Chloroethane-d5	1.529	69	134863	26.133	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery = 104.520%			
11) 1,1-Dichloroethene-d2	2.053	65	69347	24.068	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery = 96.280%			
21) 2-Butanone-d5	3.809	46	91743m	61.781	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery = 123.560%			
24) Chloroform-d	4.246	84	307969	27.517	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery = 110.080%			
26) 1,2-Dichloroethane-d4	4.941	65	166543	28.381	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery = 113.520%			
32) Benzene-d6	4.957	84	601567	26.938	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery = 107.760%			
36) 1,2-Dichloropropane-d6	5.986	67	202218	28.755	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery = 115.040%			
41) Toluene-d8	7.239	98	499092	25.330	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery = 101.320%			
43) trans-1,3-Dichloroprop...	7.551	79	75555	26.203	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery = 104.800%			
47) 2-Hexanone-d5	8.030	63	62443	65.537	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery = 131.080%			
56) 1,1,2,2-Tetrachloroeth...	10.149	84	169829	32.317	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery = 129.280%#			
66) 1,2-Dichlorobenzene-d4	11.558	152	206613	31.258	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery = 125.040%#			
Target Compounds						
16) Methylene chloride	2.445	84	38798	4.871	ug/L	99
17) trans-1,2-Dichloroethene	2.696	96	5812	0.971	ug/L	95
19) 1,1-Dichloroethane	3.111	63	14452	1.267	ug/L	99
20) cis-1,2-Dichloroethene	3.806	96	6338617	1053.406	ug/L	97
25) Chloroform	4.272	83	8491	0.724	ug/L	97
34) Trichloroethene	5.828	95	260412	34.581	ug/L	99
42) Toluene	7.330	91	13208	0.526	ug/L	93
46) Tetrachloroethene	7.902	164	699466	127.001	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061124\
Data File : VV036036.D
Acq On : 11 Jun 2024 15:56
Operator : SY/MD
Sample : P2798-06
Misc : 5.31g/10.0mL/MSVOA_V/SOIL/A
ALS Vial : 10 Sample Multiplier: 1


Instrument :
MSVOA_V
ClientSampleId :
BHB17

Quant Time: Jun 12 02:08:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM053024SMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jun 12 02:05:45 2024
Response via : Initial Calibration

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/12/2024
Supervised By :Semsettin Yesilyurt 06/12/2024





#16
Methylene chloride
Concen: 4.871 ug/L
RT: 2.445 min Scan# 41
Delta R.T. 0.003 min
Lab File: VV036036.D
Acq: 11 Jun 2024 15:56




Instrument :
MSVOA_V
ClientSampleId :
BHB17



Tgt Ion: 84 Resp: 38798
Ion Ratio Lower Upper
84 100
86 63.1 44.6 82.8
49 117.3 81.1 150.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/12/2024
Supervised By :Semsettin Yesilyurt 06/12/2024



#17
trans-1,2-Dichloroethene
Concen: 0.971 ug/L
RT: 2.696 min Scan# 515
Delta R.T. 0.007 min
Lab File: VV036036.D
Acq: 11 Jun 2024 15:56

Tgt Ion: 96 Resp: 5812
Ion Ratio Lower Upper
96 100
61 148.6 99.4 184.6
98 60.3 44.9 83.3









#19
1,1-Dichloroethane
Concen: 1.267 ug/L
RT: 3.111 min Scan# 64
Delta R.T. 0.007 min
Lab File: VV036036.D
Acq: 11 Jun 2024 15:56

Instrument :
MSVOA_V
ClientSampleId :
BHB17

Tgt Ion: 63 Resp: 14452
Ion Ratio Lower Upper
63 100
65 32.1 22.4 41.6
83 14.4 9.2 17.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/12/2024
Supervised By :Semsettin Yesilyurt 06/12/2024



#20
cis-1,2-Dichloroethene
Concen: 1053.406 ug/L
RT: 3.806 min Scan# 860
Delta R.T. -0.003 min
Lab File: VV036036.D
Acq: 11 Jun 2024 15:56


Tgt Ion: 96 Resp: 6338617
Ion Ratio Lower Upper
96 100
61 128.8 87.9 163.3
98 65.2 44.4 82.5





#25
Chloroform
Concen: 0.724 ug/L
RT: 4.272 min Scan# 1005
Delta R.T. 0.003 min
Lab File: VV036036.D
Acq: 11 Jun 2024 15:56




Instrument :
MSVOA_V
ClientSampleId :
BHB17



Tgt Ion: 83 Resp: 849
Ion Ratio Lower Upper
83 100
85 67.1 45.2 84.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 06/12/2024
Supervised By :Semsettin Yesilyurt 06/12/2024



#34
Trichloroethene
Concen: 34.581 ug/L
RT: 5.828 min Scan# 1489
Delta R.T. 0.000 min
Lab File: VV036036.D
Acq: 11 Jun 2024 15:56

Tgt Ion: 95 Resp: 260412
Ion Ratio Lower Upper
95 100
97 64.1 44.8 83.2
132 93.0 66.8 124.2
130 99.8 70.0 130.0









#42
Toluene
Concen: 0.526 ug/L
RT: 7.330 min Scan# 1950
Delta R.T. 0.019 min
Lab File: VV036036.D
Acq: 11 Jun 2024 15:56

Instrument :
MSVOA_V
ClientSampleId :
BHB17

Tgt Ion: 91 Resp: 13208
Ion Ratio Lower Upper
91 100
92 53.0 40.6 75.4

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/12/2024
Supervised By :Semsettin Yesilyurt 06/12/2024



#46
Tetrachloroethene
Concen: 127.001 ug/L
RT: 7.902 min Scan# 2134
Delta R.T. 0.000 min
Lab File: VV036036.D
Acq: 11 Jun 2024 15:56

Tgt Ion:164 Resp: 699466
Ion Ratio Lower Upper
164 100
129 94.0 64.6 120.0
131 90.5 63.2 117.4
166 127.9 89.5 166.3

