

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061219\
 Data File : VV011429.D
 Acq On : 12 Jun 2019 19:18
 Operator : SY/MD
 Sample : K3310-02
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COLZ7

Quant Time: Jun 13 11:50:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jun 13 02:32:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	139676	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	126956	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	46717	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	33998	3.21	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	64.20%
7) Chloroethane-d5	1.58	69	29853	3.39	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	67.80%
11) 1,1-Dichloroethene-d2	2.12	63	64709	3.06	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.20%
20) 2-Butanone-d5	3.97	46	116668	48.96	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.92%
24) Chloroform-d	4.40	84	85617	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.00%
26) 1,2-Dichloroethane-d4	5.08	65	49113	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.60%
32) Benzene-d6	5.09	84	165623	4.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.20%
36) 1,2-Dichloropropane-d6	6.12	67	51330	4.39	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.80%
41) Toluene-d8	7.36	98	140570	4.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.60%
43) trans-1,3-Dichloropropene-	7.66	79	14438	3.24	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	64.80%
46) 2-Hexanone-d5	8.14	63	86416	45.87	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.74%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	35266	4.42	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	43912	4.76	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061219\
Data File : VV011429.D
Acq On : 12 Jun 2019 19:18
Operator : SY/MD
Sample : K3310-02
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0LZ7

Quant Time: Jun 13 11:50:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jun 13 02:32:13 2019
Response via : Initial Calibration

