

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061219\
 Data File : VV011430.D
 Acq On : 12 Jun 2019 19:48
 Operator : SY/MD
 Sample : K3310-03
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COLZ8

Quant Time: Jun 13 11:51:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jun 13 02:32:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	134439	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	122517	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	46168	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	34518	3.39	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	67.80%
7) Chloroethane-d5	1.57	69	29325	3.46	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	69.20%
11) 1,1-Dichloroethene-d2	2.12	63	62241	3.06	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.20%
20) 2-Butanone-d5	3.96	46	120107	52.37	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.74%
24) Chloroform-d	4.40	84	83207	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.80%
26) 1,2-Dichloroethane-d4	5.08	65	48116	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.20%
32) Benzene-d6	5.09	84	162911	4.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.80%
36) 1,2-Dichloropropane-d6	6.12	67	50803	4.50	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.00%
41) Toluene-d8	7.36	98	138761	4.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.40%
43) trans-1,3-Dichloropropene-	7.66	79	14241	3.31	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	66.20%
46) 2-Hexanone-d5	8.14	63	82976	45.64	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.28%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	35523	4.61	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.20%
64) 1,2-Dichlorobenzene-d4	11.67	152	43645	4.79	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061219\
Data File : VV011430.D
Acq On : 12 Jun 2019 19:48
Operator : SY/MD
Sample : K3310-03
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COLZ8

Quant Time: Jun 13 11:51:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jun 13 02:32:13 2019
Response via : Initial Calibration

