

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061219\
 Data File : VV011441.D
 Acq On : 13 Jun 2019 01:22
 Operator : SY/MD
 Sample : K3310-15
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COM16

Quant Time: Jun 13 04:33:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jun 13 02:32:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	134147	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	123385	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	45693	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	32360	3.18	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	63.60%
7) Chloroethane-d5	1.57	69	28563	3.38	ug/L	-0.01
Spiked Amount	5.000	Range	65 - 130	Recovery	=	67.60%
11) 1,1-Dichloroethene-d2	2.12	63	55180	2.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	54.40%#
20) 2-Butanone-d5	3.96	46	114285	49.94	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.88%
24) Chloroform-d	4.40	84	77758	4.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.00%
26) 1,2-Dichloroethane-d4	5.08	65	45852	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.80%
32) Benzene-d6	5.09	84	150189	4.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.40%
36) 1,2-Dichloropropane-d6	6.12	67	48237	4.24	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.80%
41) Toluene-d8	7.36	98	129896	3.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.60%
43) trans-1,3-Dichloropropene-	7.66	79	12946	2.99	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	59.80%
46) 2-Hexanone-d5	8.14	63	77459	42.30	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.60%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	33447	4.31	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.20%
64) 1,2-Dichlorobenzene-d4	11.67	152	41870	4.64	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061219\
Data File : VV011441.D
Acq On : 13 Jun 2019 01:22
Operator : SY/MD
Sample : K3310-15
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COM16

Quant Time: Jun 13 04:33:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jun 13 02:32:13 2019
Response via : Initial Calibration

