

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061219\
 Data File : VV011449.D
 Acq On : 13 Jun 2019 07:02
 Operator : SY/MD
 Sample : K3289-12
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFC55

Quant Time: Jun 13 09:15:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jun 13 08:24:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	137555	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	125717	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	46363	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	32492	3.12	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	62.40%
7) Chloroethane-d5	1.58	69	28237	3.26	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	65.20%
11) 1,1-Dichloroethene-d2	2.12	63	59234	2.84	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	56.80%#
20) 2-Butanone-d5	3.97	46	113744	48.47	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.94%
24) Chloroform-d	4.40	84	81352	4.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.80%
26) 1,2-Dichloroethane-d4	5.08	65	47258	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.40%
32) Benzene-d6	5.09	84	156406	4.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.20%
36) 1,2-Dichloropropane-d6	6.12	67	50616	4.37	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.40%
41) Toluene-d8	7.36	98	131948	3.82	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.40%
43) trans-1,3-Dichloropropene-	7.66	79	12709	2.88	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	57.60%
46) 2-Hexanone-d5	8.14	63	80820	43.32	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.64%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	34689	4.39	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	41898	4.58	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061219\
Data File : VV011449.D
Acq On : 13 Jun 2019 07:02
Operator : SY/MD
Sample : K3289-12
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFCS5

Quant Time: Jun 13 09:15:52 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jun 13 08:24:59 2019
Response via : Initial Calibration

