

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061219\
 Data File : VV011458.D
 Acq On : 13 Jun 2019 14:06
 Operator : SY/MD
 Sample : K3314-01
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COM14

Quant Time: Jun 14 05:01:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 14 04:38:08 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	154223	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	139608	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	51566	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	30136	2.58	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	51.60%
7) Chloroethane-d5	1.58	69	28926	2.98	ug/L	0.03
Spiked Amount	5.000	Range	65 - 130	Recovery	=	59.60%#
11) 1,1-Dichloroethene-d2	2.12	63	55844	2.39	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	47.80%#
20) 2-Butanone-d5	3.97	46	117453	44.64	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.28%
24) Chloroform-d	4.40	84	78107	3.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.00%
26) 1,2-Dichloroethane-d4	5.08	65	46516	4.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.60%
32) Benzene-d6	5.10	84	150007	3.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	71.80%
36) 1,2-Dichloropropane-d6	6.12	67	48345	3.76	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	75.20%
41) Toluene-d8	7.36	98	127296	3.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	66.40%#
43) trans-1,3-Dichloropropene-	7.67	79	13696	2.79	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	55.80%
46) 2-Hexanone-d5	8.14	63	84691	40.88	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.76%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	33979	3.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	77.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	41062	4.03	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	80.60%

Target Compounds					Ovalue
34) Trichloroethene	5.96	95	5459	0.476 ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061219\
Data File : VV011458.D
Acq On : 13 Jun 2019 14:06
Operator : SY/MD
Sample : K3314-01
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COM14

Quant Time: Jun 14 05:01:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 14 04:38:08 2019
Response via : Initial Calibration

