

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061219\
 Data File : VV011459.D
 Acq On : 13 Jun 2019 14:34
 Operator : SY/MD
 Sample : K3314-02
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COM15

Quant Time: Jun 14 05:07:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 14 04:38:08 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	140250	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	126467	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	48418	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	30718	2.89	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	57.80%
7) Chloroethane-d5	1.58	69	32506	3.68	ug/L	0.03
Spiked Amount	5.000	Range	65 - 130	Recovery	=	73.60%
11) 1,1-Dichloroethene-d2	2.12	63	54218	2.55	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	51.00%#
20) 2-Butanone-d5	3.96	46	121774	50.90	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.80%
24) Chloroform-d	4.40	84	80894	4.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.60%
26) 1,2-Dichloroethane-d4	5.08	65	47072	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.20%
32) Benzene-d6	5.10	84	153177	4.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.00%
36) 1,2-Dichloropropane-d6	6.12	67	50672	4.35	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.00%
41) Toluene-d8	7.36	98	131805	3.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.80%
43) trans-1,3-Dichloropropene-	7.67	79	13663	3.08	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	61.60%
46) 2-Hexanone-d5	8.14	63	90457	48.20	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.40%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	36118	4.54	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	42332	4.43	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.60%

Target Compounds					Ovalue
13) Acetone	2.21	43	5074	3.163 ug/L	96
34) Trichloroethene	5.96	95	1498	0.144 ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061219\
Data File : VV011459.D
Acq On : 13 Jun 2019 14:34
Operator : SY/MD
Sample : K3314-02
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COM15

Quant Time: Jun 14 05:07:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 14 04:38:08 2019
Response via : Initial Calibration

