

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061219\
 Data File : VV011464.D
 Acq On : 13 Jun 2019 16:54
 Operator : SY/MD
 Sample : K3291-03
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Jun 14 05:37:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 14 04:38:08 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	132858	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	119529	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	44148	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	42901	4.26	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.20%
7) Chloroethane-d5	1.58	69	37053	4.43	ug/L	0.03
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.60%
11) 1,1-Dichloroethene-d2	2.12	63	75019	3.73	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.60%
20) 2-Butanone-d5	3.97	46	121710	53.70	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.40%
24) Chloroform-d	4.40	84	88384	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.80%
26) 1,2-Dichloroethane-d4	5.08	65	52403	5.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.40%
32) Benzene-d6	5.10	84	176718	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.80%
36) 1,2-Dichloropropane-d6	6.12	67	56440	5.12	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.40%
41) Toluene-d8	7.36	98	150544	4.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.60%
43) trans-1,3-Dichloropropene-	7.66	79	15146	3.61	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	72.20%
46) 2-Hexanone-d5	8.14	63	94555	53.31	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.62%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	37802	5.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	45020	5.17	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061219\
Data File : VV011464.D
Acq On : 13 Jun 2019 16:54
Operator : SY/MD
Sample : K3291-03
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jun 14 05:37:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 14 04:38:08 2019
Response via : Initial Calibration

