

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061219\
 Data File : VV011494.D
 Acq On : 14 Jun 2019 08:11
 Operator : SY/MD
 Sample : VIBLK40
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK40

Quant Time: Jun 14 09:27:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 14 09:13:02 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	111214	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	96970	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	33753	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	43707	5.19	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.80%
7) Chloroethane-d5	1.58	69	37688	5.38	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.60%
11) 1,1-Dichloroethene-d2	2.12	63	69300	4.12	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.40%
20) 2-Butanone-d5	3.98	46	112204	59.14	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.28%
24) Chloroform-d	4.40	84	87460	5.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	118.00%
26) 1,2-Dichloroethane-d4	5.08	65	51047	6.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	127.40%
32) Benzene-d6	5.10	84	167275	5.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	115.40%
36) 1,2-Dichloropropane-d6	6.12	67	53361	5.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	119.40%
41) Toluene-d8	7.36	98	133911	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.60%
43) trans-1,3-Dichloropropene-	7.67	79	13782	4.05	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.00%
46) 2-Hexanone-d5	8.14	63	90058	62.58	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	125.16%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	35658	5.85	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	117.00%
64) 1,2-Dichlorobenzene-d4	11.67	152	37681	5.66	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061219\
Data File : VV011494.D
Acq On : 14 Jun 2019 08:11
Operator : SY/MD
Sample : VIBLK40
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK40

Quant Time: Jun 14 09:27:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 14 09:13:02 2019
Response via : Initial Calibration

