

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061219\
 Data File : VV011421.D
 Acq On : 12 Jun 2019 14:36
 Operator : SY/MD
 Sample : K3291-03
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Jun 13 09:00:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jun 13 01:13:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	154163	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	141274	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	54196	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	42898	3.67	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.40%
7) Chloroethane-d5	1.58	69	30489	3.14	ug/L	0.02
Spiked Amount	5.000	Range	65 - 130	Recovery	=	62.80%#
11) 1,1-Dichloroethene-d2	2.12	63	69696	2.99	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	59.80%#
20) 2-Butanone-d5	3.97	46	98189	37.34	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	74.68%
24) Chloroform-d	4.40	84	78353	3.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.20%
26) 1,2-Dichloroethane-d4	5.08	65	43935	3.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.00%
32) Benzene-d6	5.09	84	157809	3.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	74.60%
36) 1,2-Dichloropropane-d6	6.12	67	47705	3.66	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	73.20%
41) Toluene-d8	7.36	98	133674	3.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	68.80%#
43) trans-1,3-Dichloropropene-	7.67	79	13483	2.72	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	54.40%#
46) 2-Hexanone-d5	8.14	63	73523	35.07	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	70.14%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	30673	3.45	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	69.00%
64) 1,2-Dichlorobenzene-d4	11.67	152	40819	3.82	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	76.40%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061219\
Data File : VV011421.D
Acq On : 12 Jun 2019 14:36
Operator : SY/MD
Sample : K3291-03
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jun 13 09:00:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jun 13 01:13:22 2019
Response via : Initial Calibration

