

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061219\
 Data File : VV011433.D
 Acq On : 12 Jun 2019 21:19
 Operator : SY/MD
 Sample : K3310-06
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COM01

Quant Time: Jun 13 11:52:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jun 13 02:32:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	141503	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	125378	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	46625	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	33829	3.16	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	63.20%
7) Chloroethane-d5	1.58	69	29699	3.33	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	66.60%
11) 1,1-Dichloroethene-d2	2.12	63	60507	2.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	56.40%#
20) 2-Butanone-d5	3.97	46	113074	46.84	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.68%
24) Chloroform-d	4.40	84	79881	4.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.60%
26) 1,2-Dichloroethane-d4	5.08	65	45864	4.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.00%
32) Benzene-d6	5.09	84	155607	4.15	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.00%
36) 1,2-Dichloropropane-d6	6.12	67	50308	4.35	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.00%
41) Toluene-d8	7.36	98	133381	3.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.40%
43) trans-1,3-Dichloropropene-	7.67	79	13720	3.12	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	62.40%
46) 2-Hexanone-d5	8.14	63	82794	44.50	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.00%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	34081	4.33	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	41928	4.56	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061219\
Data File : VV011433.D
Acq On : 12 Jun 2019 21:19
Operator : SY/MD
Sample : K3310-06
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COM01

Quant Time: Jun 13 11:52:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jun 13 02:32:13 2019
Response via : Initial Calibration

