

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061219\
 Data File : VV011438.D
 Acq On : 12 Jun 2019 23:51
 Operator : SY/MD
 Sample : K3310-12
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COM10

Quant Time: Jun 13 04:17:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jun 13 02:32:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	141636	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	128437	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	48071	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	33726	3.14	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	62.80%
7) Chloroethane-d5	1.58	69	28637	3.21	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	64.20%#
11) 1,1-Dichloroethene-d2	2.12	63	60153	2.81	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	56.20%#
20) 2-Butanone-d5	3.97	46	115300	47.72	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.44%
24) Chloroform-d	4.40	84	82967	4.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.80%
26) 1,2-Dichloroethane-d4	5.08	65	49307	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.60%
32) Benzene-d6	5.09	84	160864	4.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.80%
36) 1,2-Dichloropropane-d6	6.12	67	52142	4.41	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.20%
41) Toluene-d8	7.36	98	137770	3.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.00%
43) trans-1,3-Dichloropropene-	7.66	79	13391	2.97	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	59.40%
46) 2-Hexanone-d5	8.14	63	83826	43.98	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.96%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	34850	4.32	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	43394	4.57	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061219\
Data File : VV011438.D
Acq On : 12 Jun 2019 23:51
Operator : SY/MD
Sample : K3310-12
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COM10

Quant Time: Jun 13 04:17:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jun 13 02:32:13 2019
Response via : Initial Calibration

