

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061219\
 Data File : VV011439.D
 Acq On : 13 Jun 2019 00:21
 Operator : SY/MD
 Sample : K3310-13
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COM12

Quant Time: Jun 13 04:19:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jun 13 02:32:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	144176	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	131116	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	48758	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	34104	3.12	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	62.40%
7) Chloroethane-d5	1.58	69	28083	3.09	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	61.80%#
11) 1,1-Dichloroethene-d2	2.12	63	61539	2.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	56.40%#
20) 2-Butanone-d5	3.97	46	118888	48.34	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.68%
24) Chloroform-d	4.40	84	82877	4.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.20%
26) 1,2-Dichloroethane-d4	5.08	65	48408	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.20%
32) Benzene-d6	5.09	84	160969	4.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.00%
36) 1,2-Dichloropropane-d6	6.12	67	51709	4.28	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.60%
41) Toluene-d8	7.36	98	140417	3.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.00%
43) trans-1,3-Dichloropropene-	7.67	79	13759	2.99	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	59.80%
46) 2-Hexanone-d5	8.14	63	84603	43.48	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.96%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	36264	4.40	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.00%
64) 1,2-Dichlorobenzene-d4	11.67	152	44473	4.62	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061219\
Data File : VV011439.D
Acq On : 13 Jun 2019 00:21
Operator : SY/MD
Sample : K3310-13
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COM12

Quant Time: Jun 13 04:19:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jun 13 02:32:13 2019
Response via : Initial Calibration

