

Data File : VV011456.D
Acq On : 13 Jun 2019 12:51
Operator : SY/MD
Sample : VV0612WBL04
Misc : 25ML/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK59

Quant Time: Jun 14 04:50:58 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 14 04:38:08 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	135484	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	129455	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	50014	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	45550	4.44	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.80%
7) Chloroethane-d5	1.58	69	33360	3.91	ug/L	0.03
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.20%
11) 1,1-Dichloroethene-d2	2.11	63	74872	3.65	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.00%
20) 2-Butanone-d5	3.97	46	125032	54.10	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.20%
24) Chloroform-d	4.39	84	81724	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.40%
26) 1,2-Dichloroethane-d4	5.08	65	50400	5.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.20%
32) Benzene-d6	5.09	84	181820	4.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.80%
36) 1,2-Dichloropropane-d6	6.11	67	54687	4.58	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.60%
41) Toluene-d8	7.36	98	154451	4.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.80%
43) trans-1,3-Dichloropropene-	7.66	79	15303	3.37	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	67.40%
46) 2-Hexanone-d5	8.15	63	94946	49.42	ug/L	0.01
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.84%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	38442	4.73	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	49287	4.99	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.80%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV011456.D
Acq On : 13 Jun 2019 12:51
Operator : SY/MD
Sample : VV0612WBL04
Misc : 25ML/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK59

Quant Time: Jun 14 04:50:58 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 14 04:38:08 2019
Response via : Initial Calibration

