

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061219\
 Data File : VV011476.D
 Acq On : 13 Jun 2019 23:24
 Operator : SY/MD
 Sample : VV0612WBL05
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK60

Quant Time: Jun 14 06:39:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 14 06:06:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	120895	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	108256	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	39280	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	42045	4.59	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.80%
7) Chloroethane-d5	1.58	69	36981	4.86	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.20%
11) 1,1-Dichloroethene-d2	2.12	63	71204	3.89	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.80%
20) 2-Butanone-d5	3.97	46	121189	58.76	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	117.52%
24) Chloroform-d	4.40	84	87504	5.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.60%
26) 1,2-Dichloroethane-d4	5.08	65	51693	5.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	118.60%
32) Benzene-d6	5.10	84	172381	5.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.40%
36) 1,2-Dichloropropane-d6	6.12	67	54750	5.49	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.80%
41) Toluene-d8	7.36	98	144250	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.00%
43) trans-1,3-Dichloropropene-	7.67	79	14668	3.86	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.20%
46) 2-Hexanone-d5	8.14	63	95740	59.59	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	119.18%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	36357	5.34	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	43403	5.60	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061219\
Data File : VV011476.D
Acq On : 13 Jun 2019 23:24
Operator : SY/MD
Sample : VV0612WBL05
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK60

Quant Time: Jun 14 06:39:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 14 06:06:48 2019
Response via : Initial Calibration

