

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061324\
 Data File : VV036049.D
 Acq On : 13 Jun 2024 09:34
 Operator : SY/MD
 Sample : VV0613SBL01
 Misc : 5.00g/10.0mL/MSVOA_V/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK320

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/20/2024
 Supervised By :Semsettin Yesilyurt 06/20/2024

Quant Time: Jun 14 01:15:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM053024SMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 14 01:10:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	370388	25.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	377017	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	166924	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.272	65	121799	17.886	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	71.560%
7) Chloroethane-d5	1.529	69	112176	20.660	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.640%
11) 1,1-Dichloroethene-d2	2.053	65	58878	19.422	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	77.680%
21) 2-Butanone-d5	3.847	46	73121m	46.801	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	93.600%
24) Chloroform-d	4.243	84	256938	21.820	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	87.280%
26) 1,2-Dichloroethane-d4	4.941	65	141983	22.997	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	92.000%
32) Benzene-d6	4.953	84	481338	21.478	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	85.920%
36) 1,2-Dichloropropane-d6	5.985	67	164198	23.266	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	93.080%
41) Toluene-d8	7.239	98	385285	19.485	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	77.920%
43) trans-1,3-Dichloroprop...	7.555	79	58349	20.164	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	80.640%
47) 2-Hexanone-d5	8.034	63	38361	40.119	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.240%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	138882	26.335	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	105.320%
66) 1,2-Dichlorobenzene-d4	11.558	152	152226	24.868	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	99.480%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061324\
Data File : VV036049.D
Acq On : 13 Jun 2024 09:34
Operator : SY/MD
Sample : VV0613SBL01
Misc : 5.00g/10.0mL/MSVOA_V/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK320

Quant Time: Jun 14 01:15:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM053024SMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 14 01:10:24 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/20/2024
Supervised By :Semsettin Yesilyurt 06/20/2024

