



Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061324\
Data File : VV036047.D
Acq On : 13 Jun 2024 08:05
Operator : SY/MD
Sample : BFB321
Misc : 5.00g/10.0mL/MSVOA_V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFB321

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM053024SMA.M
Title : VOC Analysis
Last Update : Thu Jun 20 05:02:48 2024



AutoFind: Scans 2788, 2789, 2790; Background Corrected with Scan 2777

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.4	22741	PASS
75	95	30	80	50.5	62400	PASS
95	95	100	100	100.0	123485	PASS
96	95	5	9	6.8	8422	PASS
173	174	0.00	2	1.4	1461	PASS
174	95	50	120	86.4	106661	PASS
175	174	5	9	7.4	7930	PASS
176	174	95	101	95.5	101840	PASS
177	176	5	9	6.7	6815	PASS

m/z	Abundance
39.90	268.0
44.00	772.0

Instrument :
MSVOA_V
ClientSampleId :
BFB321

m/z	Abundance
35.90	1106.0
37.00	5057.0
38.00	5036.0
39.00	2286.0
40.00	251.0
41.90	106.0
42.90	115.0
44.00	1337.0
44.90	987.0
45.90	288.0
47.00	1285.0
48.00	806.0
49.00	4948.0
50.00	23144.0
51.00	6958.0
52.00	470.0
53.00	255.0
53.90	108.0
54.90	512.0
56.00	1696.0
57.00	3221.0
58.10	151.0
60.00	1094.0
61.00	6176.0
62.00	6075.0
63.00	5410.0
64.00	1221.0
65.00	1279.0
66.10	199.0
67.00	302.0
68.00	11942.0
69.00	12431.0
69.90	1026.0
72.00	670.0
73.00	5486.0
74.00	21160.0
75.00	62664.0
76.00	5438.0
76.90	909.0
78.00	529.0
78.90	5138.0
79.90	1614.0
80.90	5063.0
81.90	1230.0
83.00	139.0
85.90	221.0
87.00	5220.0
88.00	5546.0
90.80	437.0
92.00	3729.0
93.00	5688.0
94.00	14654.0
95.00	124040.0
96.00	8430.0
97.00	368.0
103.80	539.0
104.70	190.0
105.90	496.0
107.00	144.0
111.90	109.0
114.80	122.0
115.90	539.0
116.80	1044.0
117.80	477.0
118.80	849.0
127.80	383.0
128.80	186.0
129.90	507.0
130.80	182.0
134.80	241.0
136.80	164.0
140.90	1001.0
141.80	176.0
142.90	1996.0
144.00	149.0
144.90	1116.0
145.70	284.0
146.70	102.0
147.80	270.0
149.90	154.0
152.90	102.0
154.90	306.0
156.80	212.0
158.80	133.0
160.80	178.0
170.90	461.0
171.90	1294.0
172.90	1435.0
173.90	104616.0
174.90	7904.0
175.90	100216.0
176.90	7131.0

Instrument :
MSVOA_V
ClientSampleId :
BFB321

177.90

166.0

Instrument :
MSVOA_V
ClientSampleId :
BFB321

m/z	Abundance
36.00	1115.0
37.00	5192.0
38.00	5013.0
39.00	2215.0
40.00	306.0
42.00	101.0
42.90	111.0
44.00	1327.0
44.90	945.0
46.00	259.0
47.00	1172.0
48.00	766.0
49.00	5007.0
50.00	23360.0
51.00	7168.0
52.00	445.0
52.90	254.0
53.90	110.0
54.90	480.0
56.00	1684.0
57.00	3214.0
58.10	212.0
60.00	1163.0
61.00	5960.0
62.00	6290.0
63.00	5373.0
64.00	1130.0
65.00	1376.0
66.00	173.0
67.00	308.0
68.00	11960.0
69.00	12675.0
70.00	967.0
72.00	707.0
73.00	5703.0
74.00	21672.0
75.00	64448.0
76.00	5357.0
76.90	778.0
78.00	542.0
78.90	5187.0
79.90	1712.0
80.90	5134.0
81.90	1300.0
83.00	143.0
85.90	231.0
87.00	5319.0
88.00	5387.0
90.80	457.0
92.00	3768.0
93.00	5882.0
94.00	14913.0
95.00	127144.0
96.00	8666.0
97.00	373.0
103.80	526.0
104.70	206.0
105.90	551.0
107.00	138.0
108.10	103.0
112.90	124.0
114.70	140.0
115.80	499.0
116.90	990.0
117.80	487.0
118.80	851.0
127.80	364.0
128.80	202.0
129.90	556.0
130.80	209.0
134.80	275.0
136.80	191.0
140.90	1092.0
141.80	155.0
142.90	1991.0
144.00	147.0
144.90	1057.0
145.80	242.0
146.70	107.0
147.90	259.0
149.90	138.0
152.90	105.0
154.90	367.0
156.90	251.0
158.80	130.0
160.80	176.0
170.90	440.0
171.90	1400.0
172.90	1514.0
173.90	110136.0
174.90	8210.0
175.90	105144.0

Instrument :
MSVOA_V
ClientSampleId :
BFB321

176.907024.0

177.90169.0

Instrument :
MSVOA_V
ClientSampleId :
BFB321

m/z	Abundance
36.00	1016.0
37.00	4970.0
38.00	4579.0
39.00	1987.0
40.00	346.0
40.90	114.0
42.90	106.0
44.00	1221.0
44.90	870.0
46.00	231.0
47.00	1156.0
48.00	654.0
49.00	4572.0
50.00	21720.0
51.00	6737.0
52.00	436.0
52.90	287.0
53.90	107.0
54.90	398.0
56.00	1476.0
57.00	3015.0
58.10	228.0
60.00	1112.0
61.00	5398.0
62.00	5885.0
63.00	4869.0
64.00	1002.0
65.00	1326.0
65.90	132.0
67.10	297.0
68.00	11241.0
69.00	11999.0
70.00	804.0
72.00	663.0
73.00	5432.0
74.00	20240.0
75.00	60088.0
76.00	4763.0
77.10	634.0
78.00	530.0
78.90	4930.0
79.90	1578.0
80.90	4818.0
81.90	1205.0
83.00	122.0
85.90	245.0
87.00	4939.0
88.00	4856.0
90.80	416.0
92.00	3404.0
93.00	5291.0
94.00	13719.0
95.00	119272.0
96.00	8170.0
97.00	290.0
103.90	489.0
104.80	194.0
105.90	568.0
106.90	117.0
108.10	103.0
110.90	102.0
114.70	145.0
115.80	413.0
116.90	919.0
117.90	443.0
118.90	808.0
123.60	100.0
127.70	368.0
128.80	179.0
129.80	488.0
130.80	180.0
134.80	255.0
136.90	196.0
140.90	1024.0
141.80	120.0
142.80	1976.0
144.00	122.0
144.90	947.0
145.80	205.0
147.90	276.0
152.80	105.0
154.90	379.0
156.80	238.0
158.90	114.0
160.80	147.0
170.90	402.0
171.90	1364.0
172.90	1434.0
173.90	105232.0
174.90	7676.0
175.90	100160.0
176.90	6289.0

Instrument :
MSVOA_V
ClientSampleId :
BFB321

177.90

182.0

Instrument :
MSVOA_V
ClientSampleId :
BFB321