

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061522\
 Data File : VW026277.D
 Acq On : 15 Jun 2022 14:57
 Operator : SY/MD
 Sample : N3320-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGJ76

Quant Time: Jun 16 01:39:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jun 16 01:38:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	197308	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	184936	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	79367	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.317	65	80403	4.692	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.800%
7) Chloroethane-d5	1.577	69	64435	4.782	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.600%
11) 1,1-Dichloroethene-d2	2.117	63	116573	3.961	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.200%
20) 2-Butanone-d5	3.915	46	85935	51.390	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.780%
24) Chloroform-d	4.358	84	127425	5.073	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.400%
26) 1,2-Dichloroethane-d4	5.040	65	53738	5.057	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.200%
32) Benzene-d6	5.056	84	253105	5.119	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.400%
36) 1,2-Dichloropropane-d6	6.072	67	76137	5.276	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.600%
41) Toluene-d8	7.320	98	207753	4.608	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.200%
43) trans-1,3-Dichloroprop...	7.628	79	19597	4.188	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.800%
46) 2-Hexanone-d5	8.091	63	63863	45.182	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.360%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	44784	5.038	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.800%
66) 1,2-Dichlorobenzene-d4	11.625	152	65978	5.397	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.000%
Target Compounds						
5) Vinyl chloride	1.320	62	40680	2.388	ug/L	100
12) 1,1-Dichloroethene	2.127	96	13414	1.076	ug/L	# 1
13) Acetone	2.211	43	20945	16.162	ug/L	92
14) Carbon disulfide	2.304	76	6568	0.214	ug/L	98
18) trans-1,2-Dichloroethene	2.770	96	1346	0.111	ug/L	84
19) 1,1-Dichloroethane	3.204	63	10118	0.439	ug/L	98
22) cis-1,2-Dichloroethene	3.921	96	352904	28.851	ug/L	99
34) Trichloroethene	5.921	95	21988	1.699	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061522\
Data File : VV026277.D
Acq On : 15 Jun 2022 14:57
Operator : SY/MD
Sample : N3320-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGJ76

Quant Time: Jun 16 01:39:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jun 16 01:38:03 2022
Response via : Initial Calibration

