

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061522\
 Data File : VW026279.D
 Acq On : 15 Jun 2022 15:45
 Operator : SY/MD
 Sample : N3275-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BC4

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 06/17/2022
 Supervised By :Mahesh Dadoda 06/17/2022

Quant Time: Jun 16 01:40:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jun 16 01:38:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	195835	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	189319	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	79470	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.317	65	82210	4.833	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.600%
7) Chloroethane-d5	1.577	69	66713	4.988	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.800%
11) 1,1-Dichloroethene-d2	2.117	63	112874	3.864	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.200%
20) 2-Butanone-d5	3.918	46	100389	60.485	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	120.960%
24) Chloroform-d	4.358	84	129520	5.195	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.000%
26) 1,2-Dichloroethane-d4	5.040	65	55266	5.240	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.800%
32) Benzene-d6	5.056	84	257003	5.078	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.600%
36) 1,2-Dichloropropane-d6	6.075	67	78901	5.341	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.800%
41) Toluene-d8	7.316	98	209390	4.537	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.800%
43) trans-1,3-Dichloroprop...	7.628	79	20463	4.272	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.400%
46) 2-Hexanone-d5	8.091	63	72595	50.170	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.340%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	47217	5.189	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.800%
66) 1,2-Dichlorobenzene-d4	11.622	152	69570	5.683	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.600%
Target Compounds						Qvalue
9) Trichlorofluoromethane	1.764	101	11111	0.476	ug/L	99
13) Acetone	2.201	43	517645	402.444	ug/L	87
21) 2-Butanone	4.034	43	6518m	3.916	ug/L	
22) cis-1,2-Dichloroethene	3.924	96	3834	0.316	ug/L	88
34) Trichloroethene	5.927	95	4067	0.307	ug/L	94
47) Tetrachloroethene	7.982	164	2389	0.243	ug/L #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061522\
Data File : VV026279.D
Acq On : 15 Jun 2022 15:45
Operator : SY/MD
Sample : N3275-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BC4

Quant Time: Jun 16 01:40:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jun 16 01:38:03 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 06/17/2022
Supervised By :Mahesh Dadoda 06/17/2022

