

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061522\
 Data File : VW026288.D
 Acq On : 15 Jun 2022 19:21
 Operator : SY/MD
 Sample : N3275-11
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COBE3

Quant Time: Jun 16 01:42:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jun 16 01:38:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	188384	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	184497	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	73606	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.317	65	77293	4.724	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	94.400%	
7) Chloroethane-d5	1.580	69	61870	4.809	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	96.200%	
11) 1,1-Dichloroethene-d2	2.117	63	102552	3.650	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	73.000%	
20) 2-Butanone-d5	3.915	46	92640	58.024	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	116.040%	
24) Chloroform-d	4.358	84	124405	5.187	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	103.800%	
26) 1,2-Dichloroethane-d4	5.043	65	51091	5.035	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	100.800%	
32) Benzene-d6	5.059	84	238047	4.826	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.600%	
36) 1,2-Dichloropropane-d6	6.075	67	73474	5.104	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	102.000%	
41) Toluene-d8	7.320	98	191605	4.260	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	85.200%	
43) trans-1,3-Dichloroprop...	7.628	79	19079	4.087	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	81.800%	
46) 2-Hexanone-d5	8.091	63	70338	49.881	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	99.760%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	44035	4.965	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	99.400%	
66) 1,2-Dichlorobenzene-d4	11.625	152	63957	5.641	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	112.800%	
Target Compounds						
					Qvalue	
9) Trichlorofluoromethane	1.764	101	120613	5.373	ug/L	99
13) Acetone	2.204	43	75123	60.715	ug/L	89
22) cis-1,2-Dichloroethene	3.924	96	4191	0.359	ug/L	85
34) Trichloroethene	5.921	95	40203	3.113	ug/L	98
45) 1,1,2-Trichloroethane	7.853	97	537	0.066	ug/L	93
47) Tetrachloroethene	7.976	164	127602	13.340	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061522\
Data File : VV026288.D
Acq On : 15 Jun 2022 19:21
Operator : SY/MD
Sample : N3275-11
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BE3

Quant Time: Jun 16 01:42:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jun 16 01:38:03 2022
Response via : Initial Calibration

