

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061522\
 Data File : VW026289.D
 Acq On : 15 Jun 2022 19:45
 Operator : SY/MD
 Sample : N3275-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BE5

Quant Time: Jun 16 01:43:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jun 16 01:38:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	192046	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	184059	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	74495	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.317	65	74786	4.483	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.600%
7) Chloroethane-d5	1.577	69	59670	4.550	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.000%
11) 1,1-Dichloroethene-d2	2.117	63	99500	3.474	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.400%
20) 2-Butanone-d5	3.915	46	95342	58.577	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	117.160%
24) Chloroform-d	4.359	84	120353	4.923	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.400%
26) 1,2-Dichloroethane-d4	5.043	65	50414	4.874	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.400%
32) Benzene-d6	5.056	84	231226	4.699	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.000%
36) 1,2-Dichloropropane-d6	6.076	67	71382	4.970	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.400%
41) Toluene-d8	7.320	98	187487	4.178	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.600%
43) trans-1,3-Dichloroprop...	7.629	79	18422	3.956	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.200%
46) 2-Hexanone-d5	8.092	63	66364	47.175	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.340%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	43245	4.888	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.800%
66) 1,2-Dichlorobenzene-d4	11.622	152	62908	5.482	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.600%
Target Compounds						Qvalue
9) Trichlorofluoromethane	1.764	101	13279	0.580	ug/L	95
13) Acetone	2.208	43	201472	159.725	ug/L	87
34) Trichloroethene	5.924	95	9468	0.735	ug/L	97
47) Tetrachloroethene	7.979	164	15487	1.623	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061522\
Data File : VV026289.D
Acq On : 15 Jun 2022 19:45
Operator : SY/MD
Sample : N3275-13
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BE5

Quant Time: Jun 16 01:43:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jun 16 01:38:03 2022
Response via : Initial Calibration

