

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061522\
 Data File : VW026291.D
 Acq On : 15 Jun 2022 20:33
 Operator : SY/MD
 Sample : N3275-15
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BE7

Quant Time: Jun 16 01:43:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jun 16 01:38:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	191395	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	183878	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	73977	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.317	65	79566	4.786	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	95.800%	
7) Chloroethane-d5	1.580	69	65266	4.993	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	99.800%	
11) 1,1-Dichloroethene-d2	2.117	63	105929	3.711	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	74.200%	
20) 2-Butanone-d5	3.918	46	96428	59.446	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	118.900%	
24) Chloroform-d	4.362	84	130069	5.338	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	106.800%	
26) 1,2-Dichloroethane-d4	5.043	65	53079	5.149	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	103.000%	
32) Benzene-d6	5.059	84	248539	5.056	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	101.200%	
36) 1,2-Dichloropropane-d6	6.075	67	76339	5.321	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	106.400%	
41) Toluene-d8	7.320	98	199611	4.453	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	89.000%	
43) trans-1,3-Dichloroprop...	7.628	79	19399	4.170	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	83.400%	
46) 2-Hexanone-d5	8.091	63	68613	48.821	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	97.640%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	45169	5.110	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	102.200%	
66) 1,2-Dichlorobenzene-d4	11.622	152	64742	5.682	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	113.600%	
Target Compounds						
9) Trichlorofluoromethane	1.764	101	55795	2.447	ug/L	99
13) Acetone	2.211	43	97003	77.165	ug/L	91
34) Trichloroethene	5.927	95	7385	0.574	ug/L	98
47) Tetrachloroethene	7.979	164	56250	5.900	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061522\
Data File : VV026291.D
Acq On : 15 Jun 2022 20:33
Operator : SY/MD
Sample : N3275-15
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BE7

Quant Time: Jun 16 01:43:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jun 16 01:38:03 2022
Response via : Initial Calibration

