

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW061522\  
 Data File : VW026298.D  
 Acq On : 15 Jun 2022 23:44  
 Operator : SY/MD  
 Sample : N3273-10DL 10X  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 36 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 COB80DL

Manual Integrations  
 APPROVED

Reviewed By :Krupa Patel 06/17/2022  
 Supervised By :Mahesh Dadoda 06/17/2022

Quant Time: Jun 16 06:55:58 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR061522WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Thu Jun 16 04:34:39 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon           | Response | Conc   | Units     | Dev(Min) |
|-------------------------------|--------|----------------|----------|--------|-----------|----------|
| -----                         |        |                |          |        |           |          |
| Internal Standards            |        |                |          |        |           |          |
| 1) 1,4-Difluorobenzene        | 5.622  | 114            | 227767   | 5.000  | ug/L      | 0.00     |
| 28) Chlorobenzene-d5          | 8.854  | 117            | 214186   | 5.000  | ug/L      | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.249 | 152            | 86319    | 5.000  | ug/L      | 0.00     |
|                               |        |                |          |        |           |          |
| System Monitoring Compounds   |        |                |          |        |           |          |
| 4) Vinyl Chloride-d3          | 1.320  | 65             | 90738    | 4.587  | ug/L      | 0.00     |
| Spiked Amount                 | 5.000  | Range 40 - 130 | Recovery | =      | 91.800%   |          |
| 7) Chloroethane-d5            | 1.581  | 69             | 79447    | 5.107  | ug/L      | 0.00     |
| Spiked Amount                 | 5.000  | Range 65 - 130 | Recovery | =      | 102.200%  |          |
| 11) 1,1-Dichloroethene-d2     | 2.114  | 63             | 129827   | 3.822  | ug/L      | 0.00     |
| Spiked Amount                 | 5.000  | Range 60 - 125 | Recovery | =      | 76.400%   |          |
| 20) 2-Butanone-d5             | 3.918  | 46             | 141657   | 73.383 | ug/L      | 0.00     |
| Spiked Amount                 | 50.000 | Range 40 - 130 | Recovery | =      | 146.760%# |          |
| 24) Chloroform-d              | 4.359  | 84             | 161345   | 5.564  | ug/L      | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 125 | Recovery | =      | 111.200%  |          |
| 26) 1,2-Dichloroethane-d4     | 5.040  | 65             | 69622    | 5.675  | ug/L      | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 130 | Recovery | =      | 113.600%  |          |
| 32) Benzene-d6                | 5.056  | 84             | 315103   | 5.503  | ug/L      | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 125 | Recovery | =      | 110.000%  |          |
| 36) 1,2-Dichloropropane-d6    | 6.076  | 67             | 96300    | 5.762  | ug/L      | 0.00     |
| Spiked Amount                 | 5.000  | Range 60 - 140 | Recovery | =      | 115.200%  |          |
| 41) Toluene-d8                | 7.317  | 98             | 247219   | 4.735  | ug/L      | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 130 | Recovery | =      | 94.600%   |          |
| 43) trans-1,3-Dichloroprop... | 7.629  | 79             | 22488    | 4.150  | ug/L      | 0.00     |
| Spiked Amount                 | 5.000  | Range 55 - 130 | Recovery | =      | 83.000%   |          |
| 46) 2-Hexanone-d5             | 8.095  | 63             | 104222   | 63.665 | ug/L      | 0.00     |
| Spiked Amount                 | 50.000 | Range 45 - 130 | Recovery | =      | 127.340%  |          |
| 56) 1,1,2,2-Tetrachloroeth... | 10.217 | 84             | 59673    | 5.796  | ug/L      | 0.00     |
| Spiked Amount                 | 5.000  | Range 65 - 120 | Recovery | =      | 116.000%  |          |
| 66) 1,2-Dichlorobenzene-d4    | 11.625 | 152            | 79683    | 5.993  | ug/L      | 0.00     |
| Spiked Amount                 | 5.000  | Range 80 - 120 | Recovery | =      | 119.800%  |          |
|                               |        |                |          |        |           |          |
| Target Compounds              |        |                |          |        |           |          |
| 9) Trichlorofluoromethane     | 1.764  | 101            | 16010    | 0.590  | ug/L      | 95       |
| 13) Acetone                   | 2.214  | 43             | 10752m   | 7.187  | ug/L      |          |
| 16) Methylene chloride        | 2.519  | 84             | 15206    | 0.828  | ug/L      | 99       |
| 47) Tetrachloroethene         | 7.976  | 164            | 81723    | 7.359  | ug/L      | 95       |
| -----                         |        |                |          |        |           |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV061522\  
Data File : VV026298.D  
Acq On : 15 Jun 2022 23:44  
Operator : SY/MD  
Sample : N3273-10DL 10X  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 36 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
C0B80DL

Manual Integrations  
APPROVED

Reviewed By :Krupa Patel 06/17/2022  
Supervised By :Mahesh Dadoda 06/17/2022

Quant Time: Jun 16 06:55:58 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR061522WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Thu Jun 16 04:34:39 2022  
Response via : Initial Calibration

