

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061522\
 Data File : VW026306.D
 Acq On : 16 Jun 2022 02:55
 Operator : SY/MD
 Sample : N3274-05DL 1000X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0B95DL

Quant Time: Jun 16 06:58:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jun 16 04:34:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	194742	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	188984	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	76427	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	78578	4.646	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	93.000%	
7) Chloroethane-d5	1.580	69	64828	4.874	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	97.400%	
11) 1,1-Dichloroethene-d2	2.121	63	106752	3.675	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	73.600%	
20) 2-Butanone-d5	3.924	46	97980	59.365	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	118.720%	
24) Chloroform-d	4.362	84	130926	5.281	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	105.600%	
26) 1,2-Dichloroethane-d4	5.043	65	53908	5.140	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	102.800%	
32) Benzene-d6	5.059	84	247759	4.904	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.000%	
36) 1,2-Dichloropropane-d6	6.075	67	78701	5.337	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	106.800%	
41) Toluene-d8	7.320	98	203754	4.423	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	88.400%	
43) trans-1,3-Dichloroprop...	7.628	79	19729	4.126	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	82.600%	
46) 2-Hexanone-d5	8.091	63	67274	46.575	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	93.160%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	46402	5.108	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	102.200%	
66) 1,2-Dichlorobenzene-d4	11.622	152	63877	5.426	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	108.600%	
Target Compounds						
9) Trichlorofluoromethane	1.767	101	5248	0.226	ug/L	97
34) Trichloroethene	5.918	95	135226	10.223	ug/L	96
47) Tetrachloroethene	7.979	164	7862	0.802	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061522\
Data File : VV026306.D
Acq On : 16 Jun 2022 02:55
Operator : SY/MD
Sample : N3274-05DL 1000X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0B95DL

Quant Time: Jun 16 06:58:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jun 16 04:34:39 2022
Response via : Initial Calibration

