

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061522\
 Data File : VW026309.D
 Acq On : 16 Jun 2022 04:07
 Operator : SY/MD
 Sample : N3274-14 10X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BB7

Quant Time: Jun 16 06:58:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jun 16 04:34:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	205418	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	200759	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	82301	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	76777	4.303	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	86.000%	
7) Chloroethane-d5	1.581	69	64256	4.580	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	91.600%	
11) 1,1-Dichloroethene-d2	2.121	63	105831	3.454	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	69.000%	
20) 2-Butanone-d5	3.921	46	88815	51.015	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	102.040%	
24) Chloroform-d	4.362	84	129598	4.956	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.200%	
26) 1,2-Dichloroethane-d4	5.044	65	52852	4.777	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	95.600%	
32) Benzene-d6	5.060	84	246803	4.598	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.000%	
36) 1,2-Dichloropropane-d6	6.076	67	75987	4.851	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	97.000%	
41) Toluene-d8	7.320	98	198876	4.064	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	81.200%	
43) trans-1,3-Dichloroprop...	7.629	79	18678	3.677	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	73.600%	
46) 2-Hexanone-d5	8.092	63	68492	44.637	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	89.280%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	44788	4.641	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	92.800%	
66) 1,2-Dichlorobenzene-d4	11.622	152	64375	5.078	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	101.600%	
Target Compounds						
					Qvalue	
9) Trichlorofluoromethane	1.767	101	64818	2.648	ug/L	99
31) Carbon tetrachloride	4.844	117	1230	0.065	ug/L	94
34) Trichloroethene	5.921	95	106766	7.598	ug/L	98
47) Tetrachloroethene	7.976	164	105676	10.153	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061522\
Data File : VV026309.D
Acq On : 16 Jun 2022 04:07
Operator : SY/MD
Sample : N3274-14 10X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BB7

Quant Time: Jun 16 06:58:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jun 16 04:34:39 2022
Response via : Initial Calibration

