

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061522\
 Data File : VW026310.D
 Acq On : 16 Jun 2022 04:31
 Operator : SY/MD
 Sample : N3274-15
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BB8

Quant Time: Jun 16 06:59:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jun 16 04:34:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	199974	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	194682	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	77930	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	74213	4.273	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.400%
7) Chloroethane-d5	1.581	69	63828	4.674	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.400%
11) 1,1-Dichloroethene-d2	2.121	63	103972	3.486	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.800%
20) 2-Butanone-d5	3.921	46	89149	52.601	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.200%
24) Chloroform-d	4.359	84	126789	4.980	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.600%
26) 1,2-Dichloroethane-d4	5.043	65	52057	4.833	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.600%
32) Benzene-d6	5.060	84	244067	4.689	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.800%
36) 1,2-Dichloropropane-d6	6.076	67	76353	5.026	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.600%
41) Toluene-d8	7.320	98	195322	4.116	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.400%
43) trans-1,3-Dichloroprop...	7.629	79	18295	3.714	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.200%
46) 2-Hexanone-d5	8.092	63	67017	45.039	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.080%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	43125	4.608	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.200%
66) 1,2-Dichlorobenzene-d4	11.622	152	63417	5.283	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.600%
Target Compounds						
					Qvalue	
9) Trichlorofluoromethane	1.767	101	113754	4.774	ug/L	98
13) Acetone	2.211	43	255104	194.226	ug/L	88
22) cis-1,2-Dichloroethene	3.934	96	1825	0.147	ug/L	75
31) Carbon tetrachloride	4.841	117	2398	0.132	ug/L	91
34) Trichloroethene	5.918	95	91818	6.738	ug/L	99
47) Tetrachloroethene	7.976	164	80770	8.002	ug/L	99
53) m,p-Xylene	9.156	106	1255	0.061	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061522\
Data File : VV026310.D
Acq On : 16 Jun 2022 04:31
Operator : SY/MD
Sample : N3274-15
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BB8

Quant Time: Jun 16 06:59:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jun 16 04:34:39 2022
Response via : Initial Calibration

