

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061522\
 Data File : VW026311.D
 Acq On : 16 Jun 2022 04:54
 Operator : SY/MD
 Sample : N3274-16 10X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BB9

Quant Time: Jun 16 06:59:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jun 16 04:34:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	214064	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	201005	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	81927	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	76607	4.120	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.400%
7) Chloroethane-d5	1.580	69	63216	4.324	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.400%
11) 1,1-Dichloroethene-d2	2.114	63	108684	3.404	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.000%
20) 2-Butanone-d5	3.918	46	80395	44.314	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	88.620%
24) Chloroform-d	4.358	84	129405	4.749	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.000%
26) 1,2-Dichloroethane-d4	5.040	65	53824	4.668	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.400%
32) Benzene-d6	5.056	84	256763	4.778	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.600%
36) 1,2-Dichloropropane-d6	6.072	67	78635	5.014	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.200%
41) Toluene-d8	7.317	98	206931	4.223	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.400%
43) trans-1,3-Dichloroprop...	7.628	79	18960	3.728	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.600%
46) 2-Hexanone-d5	8.095	63	70509	45.896	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.800%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	45512	4.711	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.200%
66) 1,2-Dichlorobenzene-d4	11.622	152	66303	5.254	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.000%
Target Compounds						
					Qvalue	
9) Trichlorofluoromethane	1.764	101	59989	2.352	ug/L	99
16) Methylene chloride	2.516	84	5279	0.306	ug/L	94
31) Carbon tetrachloride	4.844	117	1347	0.072	ug/L	97
34) Trichloroethene	5.918	95	166215	11.814	ug/L	98
47) Tetrachloroethene	7.976	164	42327	4.062	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061522\
Data File : VV026311.D
Acq On : 16 Jun 2022 04:54
Operator : SY/MD
Sample : N3274-16 10X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BB9

Quant Time: Jun 16 06:59:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jun 16 04:34:39 2022
Response via : Initial Calibration

