

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061522\
 Data File : VW026317.D
 Acq On : 16 Jun 2022 07:17
 Operator : SY/MD
 Sample : N3274-09
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0B99

Quant Time: Jun 16 08:57:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jun 16 04:34:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	201925	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	191437	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	76435	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	75253	4.291	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	85.800%	
7) Chloroethane-d5	1.581	69	63537	4.607	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	92.200%	
11) 1,1-Dichloroethene-d2	2.121	63	104775	3.479	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	69.600%	
20) 2-Butanone-d5	3.921	46	98804	57.734	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	115.460%	
24) Chloroform-d	4.362	84	128633	5.004	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	100.000%	
26) 1,2-Dichloroethane-d4	5.043	65	52279	4.807	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	96.200%	
32) Benzene-d6	5.060	84	248300	4.851	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.000%	
36) 1,2-Dichloropropane-d6	6.076	67	78508	5.256	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	105.200%	
41) Toluene-d8	7.320	98	199808	4.281	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	85.600%	
43) trans-1,3-Dichloroprop...	7.629	79	19097	3.943	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	78.800%	
46) 2-Hexanone-d5	8.092	63	74688	51.046	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	102.100%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	44281	4.812	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	96.200%	
66) 1,2-Dichlorobenzene-d4	11.622	152	65675	5.578	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	111.600%	
Target Compounds						
9) Trichlorofluoromethane	1.767	101	153386	6.375	ug/L	100
13) Acetone	2.217	43	177130	133.557	ug/L	88
22) cis-1,2-Dichloroethene	3.925	96	4274	0.341	ug/L	88
31) Carbon tetrachloride	4.844	117	4450	0.248	ug/L	99
34) Trichloroethene	5.918	95	463517	34.592	ug/L	97
47) Tetrachloroethene	7.976	164	683088	68.825	ug/L	99
51) Chlorobenzene	8.883	112	3074	0.091	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061522\
Data File : VV026317.D
Acq On : 16 Jun 2022 07:17
Operator : SY/MD
Sample : N3274-09
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0B99

Quant Time: Jun 16 08:57:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jun 16 04:34:39 2022
Response via : Initial Calibration

