

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061522\
 Data File : VW026278.D
 Acq On : 15 Jun 2022 15:21
 Operator : SY/MD
 Sample : VIBLK257
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK257

Quant Time: Jun 16 01:40:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jun 16 01:38:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	188122	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	181022	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	74709	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.317	65	85214	5.215	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 104.400%		
7) Chloroethane-d5	1.577	69	69006	5.371	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 107.400%		
11) 1,1-Dichloroethene-d2	2.118	63	117581	4.190	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 83.800%		
20) 2-Butanone-d5	3.908	46	87910	55.138	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 110.280%		
24) Chloroform-d	4.355	84	135730	5.667	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 113.400%		
26) 1,2-Dichloroethane-d4	5.040	65	55553	5.483	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 109.600%		
32) Benzene-d6	5.056	84	261733	5.408	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 108.200%		
36) 1,2-Dichloropropane-d6	6.076	67	78669	5.570	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 111.400%		
41) Toluene-d8	7.320	98	215404	4.881	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 97.600%		
43) trans-1,3-Dichloroprop...	7.625	79	20250	4.421	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 88.400%		
46) 2-Hexanone-d5	8.092	63	73384	53.040	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 106.080%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	47794	5.493	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 109.800%		
66) 1,2-Dichlorobenzene-d4	11.622	152	69474	6.037	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 120.800%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061522\
Data File : VV026278.D
Acq On : 15 Jun 2022 15:21
Operator : SY/MD
Sample : VIBLK257
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK257

Quant Time: Jun 16 01:40:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jun 16 01:38:03 2022
Response via : Initial Calibration

