

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061618\
 Data File : VV006331.D
 Acq On : 16 Jun 2018 19:58
 Operator : SY/MD
 Sample : VSTDCCC005EC
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTD00569

Quant Time: Jun 18 01:19:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Jun 18 01:15:51 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	1.000	1.000	0.0	104	0.00
2 T	Dichlorodifluoromethane	0.381	0.310	18.6	83	0.00
3 T	Chloromethane	0.319	0.322	-0.9	105	0.00
4 S	Vinyl Chloride-d3	0.273	0.328	-20.1	124	0.00
5 T	Vinyl chloride	0.386	0.364	5.7	96	0.00
6 T	Bromomethane	0.205	0.193	5.9	98	0.00
7 S	Chloroethane-d5	0.219	0.251	-14.6	120	0.00
8 T	Chloroethane	0.216	0.220	-1.9	103	0.00
9 T	Trichlorofluoromethane	0.500	0.496	0.8	99	0.00
10 T	1,1,2-Trichloro-1,2,2-trifl	0.309	0.305	1.3	99	0.00
11 S	1,1-Dichloroethene-d2	0.531	0.585	-10.2	112	0.00
12 T	1,1-Dichloroethene	0.287	0.287	0.0	101	0.00
13 T	Acetone	0.031	0.039	-25.8	139	-0.02
14 T	Carbon disulfide	0.961	0.653	32.0#	68	0.00
15 T	Methyl Acetate	0.094	0.109	-16.0	133	0.00
16 T	Methylene chloride	0.317	0.313	1.3	108	0.00
17 T	Methyl tert-butyl Ether	0.710	0.721	-1.5	101	0.00
18 T	trans-1,2-Dichloroethene	0.313	0.310	1.0	98	0.00
19 T	1,1-Dichloroethane	0.527	0.543	-3.0	104	0.00
20 S	2-Butanone-d5	0.058	0.062	-6.9	122	-0.01
21 T	2-Butanone	0.057	0.063	-10.5	118	-0.01
22 T	cis-1,2-Dichloroethene	0.339	0.337	0.6	100	0.00
23 T	Bromochloromethane	0.136	0.137	-0.7	100	0.00
24 S	Chloroform-d	0.529	0.545	-3.0	106	0.00
25 T	Chloroform	0.544	0.557	-2.4	104	0.00
26 S	1,2-Dichloroethane-d4	0.256	0.264	-3.1	107	0.00
27 T	1,2-Dichloroethane	0.312	0.323	-3.5	104	0.00
28 I	Chlorobenzene-d5	1.000	1.000	0.0	107	0.00
29 T	1,1,1-Trichloroethane	0.552	0.503	8.9	93	0.00
30 T	Cyclohexane	0.568	0.492	13.4	89	0.00
31 T	Carbon tetrachloride	0.487	0.415	14.8	88	0.00
32 S	Benzene-d6	1.202	1.272	-5.8	113	0.00
33 T	Benzene	1.379	1.368	0.8	102	0.00
34 T	Trichloroethene	0.376	0.370	1.6	101	0.00
35 T	Methylcyclohexane	0.647	0.558	13.8	89	0.00
36 S	1,2-Dichloropropane-d6	0.361	0.372	-3.0	110	0.00
37 T	1,2-Dichloropropane	0.330	0.327	0.9	102	0.00
38 T	Bromodichloromethane	0.440	0.371	15.7	85	0.00
39 T	cis-1,3-Dichloropropene	0.538	0.461	14.3	88	0.00
40 T	4-Methyl-2-pentanone	0.171	0.169	1.2	101	0.00
41 S	Toluene-d8	1.146	1.230	-7.3	116	0.00
42 T	Toluene	1.522	1.476	3.0	100	0.00
43 S	trans-1,3-Dichloropropene-d	0.162	0.129	20.4	87	0.00
44 T	trans-1,3-Dichloropropene	0.437	0.355	18.8	84	0.00
45 T	1,1,2-Trichloroethane	0.234	0.239	-2.1	106	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061618\
 Data File : VV006331.D
 Acq On : 16 Jun 2018 19:58
 Operator : SY/MD
 Sample : VSTDCCC005EC
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTD00569

Quant Time: Jun 18 01:19:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Jun 18 01:15:51 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 S	2-Hexanone-d5	0.070	0.061	12.9	91	0.00
47 T	Tetrachloroethene	0.301	0.298	1.0	100	0.00
48 T	2-Hexanone	0.120	0.122	-1.7	103	0.00
49 T	Dibromochloromethane	0.294	0.240	18.4	85	0.00
50 T	1,2-Dibromoethane	0.218	0.217	0.5	102	0.00
51 T	Chlorobenzene	0.977	0.983	-0.6	103	0.00
52 T	Ethylbenzene	1.695	1.636	3.5	99	0.00
53 T	m,p-xylene	0.661	0.627	5.1	98	0.00
54 T	o-xylene	0.647	0.631	2.5	99	0.00
55 T	Styrene	1.064	1.033	2.9	99	0.00
56 T	Isopropylbenzene	1.703	1.649	3.2	99	0.00
57 S	1,1,2,2-Tetrachloroethane-d	0.262	0.249	5.0	101	0.00
58 T	1,1,2,2-Tetrachloroethane	0.268	0.245	8.6	95	0.00
59	1,2,3-Trichloropropane	0.186	0.191	-2.7	104	0.00
60 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
61 T	Bromoform	0.339	0.238	29.8	73	0.00
62 T	1,3-Dichlorobenzene	1.575	1.561	0.9	103	0.00
63 T	1,4-Dichlorobenzene	1.558	1.550	0.5	104	0.00
64 S	1,2-Dichlorobenzene-d4	0.867	0.906	-4.5	113	0.00
65 T	1,2-Dichlorobenzene	1.453	1.481	-1.9	105	0.00
66 T	1,2-Dibromo-3-chloropropane	0.096	0.063	34.4#	67	0.00
67	1,3,5-Trichlorobenzene	1.212	1.200	1.0	102	0.00
68 T	1,2,4-trichlorobenzene	0.938	0.927	1.2	99	0.00
69	Naphthalene	1.577	1.393	11.7	87	0.00
70 T	1,2,3-Trichlorobenzene	0.861	0.867	-0.7	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0