

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061719\
 Data File : VV011561.D
 Acq On : 17 Jun 2019 20:05
 Operator : SY/MD
 Sample : VIBLK22
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK22

Quant Time: Jun 18 09:30:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jun 18 08:42:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	192093	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	176394	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	67526	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	45160	3.12	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	62.40%
7) Chloroethane-d5	1.58	69	38499	4.19	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.80%
11) 1,1-Dichloroethene-d2	2.12	63	75554	2.74	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	54.80%#
20) 2-Butanone-d5	3.96	46	184879	45.30	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.60%
24) Chloroform-d	4.40	84	111860	4.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.20%
26) 1,2-Dichloroethane-d4	5.08	65	66228	4.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.00%
32) Benzene-d6	5.09	84	220820	4.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.80%
36) 1,2-Dichloropropane-d6	6.12	67	70692	4.23	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.60%
41) Toluene-d8	7.36	98	186390	3.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.40%
43) trans-1,3-Dichloropropene-	7.66	79	21102	3.80	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.00%
46) 2-Hexanone-d5	8.14	63	143660	44.78	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.56%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	51472	4.11	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.20%
64) 1,2-Dichlorobenzene-d4	11.67	152	64742	4.81	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061719\
Data File : VV011561.D
Acq On : 17 Jun 2019 20:05
Operator : SY/MD
Sample : VIBLK22
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK22

Quant Time: Jun 18 09:30:01 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jun 18 08:42:07 2019
Response via : Initial Calibration

