

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061818\
 Data File : VV006334.D
 Acq On : 18 Jun 2018 11:46
 Operator : SY/MD
 Sample : VV0618WBL01
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK77

Quant Time: Jun 18 12:53:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Jun 18 03:14:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	333746	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	300370	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	135204	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	104569	5.74	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	114.80%
7) Chloroethane-d5	1.58	69	85265	5.84	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	116.80%
11) 1,1-Dichloroethene-d2	2.13	63	143416	4.05	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.00%
20) 2-Butanone-d5	3.98	46	208823	53.95	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.90%
24) Chloroform-d	4.40	84	172456	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.80%
26) 1,2-Dichloroethane-d4	5.09	65	86934	5.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.00%
32) Benzene-d6	5.10	84	381729	5.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.80%
36) 1,2-Dichloropropane-d6	6.12	67	111819	5.15	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.00%
41) Toluene-d8	7.36	98	356592	5.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.60%
43) trans-1,3-Dichloropropene-	7.67	79	39250	4.04	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.80%
46) 2-Hexanone-d5	8.14	63	166105	39.62	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	79.24%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	66804	4.25	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	123758	5.28	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061818\
Data File : VV006334.D
Acq On : 18 Jun 2018 11:46
Operator : SY/MD
Sample : VV0618WBL01
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK77

Quant Time: Jun 18 12:53:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jun 18 03:14:32 2018
Response via : Initial Calibration

