

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061819\
 Data File : VV011566.D
 Acq On : 18 Jun 2019 10:55
 Operator : SY/MD
 Sample : K3313-06
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COLY6

Quant Time: Jun 18 12:13:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jun 18 11:44:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	155752	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	149086	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	60765	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	54427	4.63	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.60%
7) Chloroethane-d5	1.58	69	39040	5.24	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.80%
11) 1,1-Dichloroethene-d2	2.12	63	97774	4.37	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.40%
20) 2-Butanone-d5	3.96	46	166896	50.44	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.88%
24) Chloroform-d	4.39	84	108146	4.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.60%
26) 1,2-Dichloroethane-d4	5.08	65	65961	5.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.60%
32) Benzene-d6	5.09	84	218566	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.20%
36) 1,2-Dichloropropane-d6	6.11	67	71084	5.03	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.60%
41) Toluene-d8	7.36	98	188075	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.80%
43) trans-1,3-Dichloropropene-	7.66	79	22667	4.82	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.40%
46) 2-Hexanone-d5	8.14	63	122197	45.07	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.14%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	47341	4.47	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	59867	4.94	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.80%

Target Compounds

					Ovalue
12) 1,1-Dichloroethene	2.13	96	26257	2.711 ug/L #	56
19) 1,1-Dichloroethane	3.23	63	2927	0.129 ug/L	97
22) cis-1,2-Dichloroethene	3.96	96	34245	2.760 ug/L	92
29) 1,1,1-Trichloroethane	4.65	97	18479	0.996 ug/L	98
34) Trichloroethene	5.95	95	484394	39.007 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061819\
Data File : VV011566.D
Acq On : 18 Jun 2019 10:55
Operator : SY/MD
Sample : K3313-06
Misc : 25ML/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COLY6

Quant Time: Jun 18 12:13:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jun 18 11:44:39 2019
Response via : Initial Calibration

