

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061819\
 Data File : VV011567.D
 Acq On : 18 Jun 2019 11:26
 Operator : SY/MD
 Sample : VIBLK23
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK23

Quant Time: Jun 18 12:14:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jun 18 11:44:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	173243	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	157861	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	62251	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	60469	4.63	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.60%
7) Chloroethane-d5	1.58	69	47359	5.71	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	114.20%
11) 1,1-Dichloroethene-d2	2.12	63	93911	3.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.40%
20) 2-Butanone-d5	3.96	46	176989	48.09	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.18%
24) Chloroform-d	4.40	84	124311	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.80%
26) 1,2-Dichloroethane-d4	5.08	65	72548	5.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.40%
32) Benzene-d6	5.09	84	250271	5.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.20%
36) 1,2-Dichloropropane-d6	6.11	67	78674	5.26	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.20%
41) Toluene-d8	7.36	98	214646	5.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.20%
43) trans-1,3-Dichloropropene-	7.66	79	23200	4.66	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.20%
46) 2-Hexanone-d5	8.14	63	134212	46.75	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.50%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	50797	4.53	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	69217	5.57	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061819\
Data File : VV011567.D
Acq On : 18 Jun 2019 11:26
Operator : SY/MD
Sample : VIBLK23
Misc : 25ML/MSVOA V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK23

Quant Time: Jun 18 12:14:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR061419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jun 18 11:44:39 2019
Response via : Initial Calibration

