

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061822\
 Data File : VW026416.D
 Acq On : 18 Jun 2022 11:37
 Operator : SY/MD
 Sample : N3278-21
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Jun 20 01:27:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jun 20 01:25:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	190735	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	179626	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	72713	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.323	65	75531	4.559	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.200%
7) Chloroethane-d5	1.584	69	63655	4.887	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.800%
11) 1,1-Dichloroethene-d2	2.124	63	110981	3.901	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.000%
20) 2-Butanone-d5	3.931	46	76847	47.539	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.080%
24) Chloroform-d	4.365	84	131650	5.422	ug/L	0.01
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.400%
26) 1,2-Dichloroethane-d4	5.043	65	53156	5.174	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.400%
32) Benzene-d6	5.063	84	260961	5.434	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.600%
36) 1,2-Dichloropropane-d6	6.076	67	77580	5.535	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.800%
41) Toluene-d8	7.320	98	214694	4.903	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.000%
43) trans-1,3-Dichloroprop...	7.629	79	17379	3.824	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.400%
46) 2-Hexanone-d5	8.095	63	60432	44.018	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.040%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	47102	5.455	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.200%
66) 1,2-Dichlorobenzene-d4	11.625	152	64319	5.743	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061822\
Data File : VV026416.D
Acq On : 18 Jun 2022 11:37
Operator : SY/MD
Sample : N3278-21
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Jun 20 01:27:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jun 20 01:25:22 2022
Response via : Initial Calibration

