




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061822\
 Data File : VV026421.D
 Acq On : 18 Jun 2022 13:35
 Operator : SY/MD
 Sample : N3345-02
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AK5

Manual IntegrationsAPPROVED

Reviewed By :Krupa Patel 06/20/2022
 Supervised By :Mahesh Dadoda 06/20/2022

Quant Time: Jun 20 02:13:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jun 20 01:25:22 2022
 Response via : Initial Calibration



TIC: VV026421.D\data.ms

(13) Acetone (T)

2.227min (+ 0.016) 1.74 ug/L

response 2281

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	30.00	4.21
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061822\
 Data File : VV026421.D
 Acq On : 18 Jun 2022 13:35
 Operator : SY/MD
 Sample : N3345-02
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AK5

Manual IntegrationsAPPROVED

Quant Time: Jun 20 01:28:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jun 20 01:25:22 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 06/20/2022
 Supervised By :Mahesh Dadoda 06/20/2022



TIC: VV026421.D\data.ms

(13) Acetone (T)

2.227min (+ 0.016) 5.68 ug/L m

response 7429

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	30.00	0.00
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061822\
 Data File : VV026421.D
 Acq On : 18 Jun 2022 13:35
 Operator : SY/MD
 Sample : N3345-02
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AK5

Manual IntegrationsAPPROVED

Quant Time: Jun 20 01:28:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jun 20 01:25:22 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 06/20/2022
 Supervised By :Mahesh Dadoda 06/20/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	199295	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	188385	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	78439	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.323	65	77241	4.462	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	89.200%	
7) Chloroethane-d5	1.584	69	65608	4.820	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	96.400%	
11) 1,1-Dichloroethene-d2	2.124	63	112566	3.787	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	75.800%	
20) 2-Butanone-d5	3.931	46	86938	51.471	ug/L	0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	102.940%	
24) Chloroform-d	4.362	84	139950	5.516	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	110.400%	
26) 1,2-Dichloroethane-d4	5.043	65	56522	5.266	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	105.400%	
32) Benzene-d6	5.059	84	276210	5.484	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	109.600%	
36) 1,2-Dichloropropane-d6	6.075	67	80371	5.468	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	109.400%	
41) Toluene-d8	7.320	98	239271	5.210	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	104.200%	
43) trans-1,3-Dichloroprop...	7.628	79	17662	3.705	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	74.200%	
46) 2-Hexanone-d5	8.091	63	69166	48.038	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	96.080%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	49457	5.462	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	109.200%	
66) 1,2-Dichlorobenzene-d4	11.625	152	69254	5.732	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	114.600%	
Target Compounds						
					Qvalue	
13) Acetone	2.227	43	7429m	5.675	ug/L	
14) Carbon disulfide	2.310	76	11411	0.368	ug/L	98
25) Chloroform	4.391	83	59952	2.373	ug/L	96
33) Benzene	5.108	78	2287	0.047	ug/L	100
42) Toluene	7.400	91	10182	0.200	ug/L	100
53) m,p-Xylene	9.149	106	1147	0.058	ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061822\
Data File : VV026421.D
Acq On : 18 Jun 2022 13:35
Operator : SY/MD
Sample : N3345-02
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AK5

Manual IntegrationsAPPROVED

Quant Time: Jun 20 01:28:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jun 20 01:25:22 2022
Response via : Initial Calibration

Reviewed By :Krupa Patel 06/20/2022
Supervised By :Mahesh Dadoda 06/20/2022

