

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061822\
 Data File : VV026414.D
 Acq On : 18 Jun 2022 10:48
 Operator : SY/MD
 Sample : VV0618WBL01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK379

Quant Time: Jun 20 01:27:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jun 20 01:25:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	196228	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	190733	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	78004	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.323	65	74306	4.360	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.200%
7) Chloroethane-d5	1.584	69	60762	4.534	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.600%
11) 1,1-Dichloroethene-d2	2.124	63	103620	3.540	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.800%
20) 2-Butanone-d5	3.928	46	76369	45.921	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.840%
24) Chloroform-d	4.362	84	125508	5.024	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.400%
26) 1,2-Dichloroethane-d4	5.040	65	52488	4.966	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.400%
32) Benzene-d6	5.056	84	242557	4.757	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.200%
36) 1,2-Dichloropropane-d6	6.072	67	74783	5.025	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.400%
41) Toluene-d8	7.317	98	201403	4.332	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.600%
43) trans-1,3-Dichloroprop...	7.625	79	15406	3.192	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	63.800%
46) 2-Hexanone-d5	8.092	63	59834	41.045	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.080%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	45647	4.979	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.600%
66) 1,2-Dichlorobenzene-d4	11.622	152	63189	5.259	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061822\
Data File : VV026414.D
Acq On : 18 Jun 2022 10:48
Operator : SY/MD
Sample : VV0618WBL01
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK379

Quant Time: Jun 20 01:27:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jun 20 01:25:22 2022
Response via : Initial Calibration

