

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061824\
 Data File : VV036112.D
 Acq On : 18 Jun 2024 11:55
 Operator : SY/MD
 Sample : P2909-02RE
 Misc : 4.98g/10.0mL/MSVOA_V/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A4B76RE

Quant Time: Jun 19 03:28:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM053024SMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jun 19 03:24:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	412427	25.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	283897	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	65537	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.272	65	158292	20.876	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.520%
7) Chloroethane-d5	1.526	69	131383	21.731	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.920%
11) 1,1-Dichloroethene-d2	2.050	65	67067	19.869	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	79.480%
21) 2-Butanone-d5	3.812	46	31625	18.178	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	36.360%
24) Chloroform-d	4.243	84	237993	18.151	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	72.600%
26) 1,2-Dichloroethane-d4	4.937	65	112196	16.320	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	65.280%#
32) Benzene-d6	4.953	84	440568	26.107	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	104.440%
36) 1,2-Dichloropropane-d6	5.986	67	138322	26.028	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	104.120%
41) Toluene-d8	7.243	98	303846	20.406	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	81.640%
43) trans-1,3-Dichloroprop...	7.555	79	39766	18.250	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	73.000%
47) 2-Hexanone-d5	8.034	63	18398	25.552	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	51.100%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	63627	16.022	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	64.080%
66) 1,2-Dichlorobenzene-d4	11.561	152	43647	18.161	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	72.640%#
Target Compounds						
					Qvalue	
13) Acetone	2.108	43	26620	9.823	ug/L #	43
14) Carbon disulfide	2.233	76	37754	1.646	ug/L	100
16) Methylene chloride	2.439	84	45879	4.916	ug/L	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061824\
Data File : VV036112.D
Acq On : 18 Jun 2024 11:55
Operator : SY/MD
Sample : P2909-02RE
Misc : 4.98g/10.0mL/MSVOA_V/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A4B76RE

Quant Time: Jun 19 03:28:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM053024SMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jun 19 03:24:16 2024
Response via : Initial Calibration





#13

Acetone

Concen: 9.823 ug/L

RT: 2.108 min Scan# 31

Delta R.T. -0.035 min

Lab File: VV036112.D

Acq: 18 Jun 2024 11:55

Instrument :

MSVOA_V

ClientSampleId :





A4B76RE

Tgt Ion: 43 Resp: 26620

Ion Ratio Lower Upper

43 100

58 34.4 0.0 24.2#



#14

Carbon disulfide

Concen: 1.646 ug/L

RT: 2.233 min Scan# 371

Delta R.T. 0.006 min

Lab File: VV036112.D




Acq: 18 Jun 2024 11:55


Tgt Ion: 76 Resp: 37754

Ion Ratio Lower Upper

76 100

78 9.2 7.4 11.2





#16
Methylene chloride
Concen: 4.916 ug/L
RT: 2.439 min Scan# 41
Delta R.T. 0.003 min
Lab File: VV036112.D
Acq: 18 Jun 2024 11:55

Instrument :
MSVOA_V
ClientSampleId :
A4B76RE

Tgt Ion:	84	Resp:	45879
Ion	Ratio	Lower	Upper
84	100		
86	61.6	44.6	82.8
49	117.2	81.1	150.5

