

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW062124\
 Data File : VW036204.D
 Acq On : 21 Jun 2024 14:09
 Operator : SY/MD
 Sample : VIBLK299
 Misc : 5.00g/10.0mL/MSVOA_V/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK299

Quant Time: Jun 22 00:13:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM053024SMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 20 05:02:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	301410	25.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	300558	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	131184	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.272	65	118759	21.431	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.720%
7) Chloroethane-d5	1.526	69	104163	23.574	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	94.280%
11) 1,1-Dichloroethene-d2	2.050	65	57403	23.269	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	93.080%
21) 2-Butanone-d5	3.821	46	71722	56.412	ug/L	-0.02
Spiked Amount	50.000	Range	20 - 135	Recovery	=	112.820%
24) Chloroform-d	4.243	84	235500	24.576	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	98.320%
26) 1,2-Dichloroethane-d4	4.937	65	137203	27.309	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	109.240%
32) Benzene-d6	4.953	84	457986	25.635	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	102.520%
36) 1,2-Dichloropropane-d6	5.985	67	150710	26.787	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	107.160%
41) Toluene-d8	7.239	98	385517	24.456	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	97.840%
43) trans-1,3-Dichloroprop...	7.554	79	58540	25.377	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	101.520%
47) 2-Hexanone-d5	8.030	63	48016	62.991	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	125.980%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	125879	29.941	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	119.760%
66) 1,2-Dichlorobenzene-d4	11.557	152	138887	28.871	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	115.480%
Target Compounds						Qvalue
16) Methylene chloride	2.439	84	40011	5.866	ug/L	96
64) 1,3-Dichlorobenzene	11.123	146	5677	0.642	ug/L	96
65) 1,4-Dichlorobenzene	11.210	146	4581	0.502	ug/L	93
70) 1,2,4-trichlorobenzene	13.207	180	2842	0.546	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV062124\
Data File : VV036204.D
Acq On : 21 Jun 2024 14:09
Operator : SY/MD
Sample : VIBLK299
Misc : 5.00g/10.0mL/MSVOA_V/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK299

Quant Time: Jun 22 00:13:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM053024SMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 20 05:02:48 2024
Response via : Initial Calibration

