

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062218\
 Data File : VV006418.D
 Acq On : 22 Jun 2018 17:34
 Operator : SY/MD
 Sample : J3615-02 10X
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BEXT8DL

Quant Time: Jun 23 00:20:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 22 23:56:13 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	315401	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	282422	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	124408	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	69599	4.04	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.80%
7) Chloroethane-d5	1.59	69	65496	4.75	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.00%
11) 1,1-Dichloroethene-d2	2.13	63	98338	2.94	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	58.80%#
20) 2-Butanone-d5	3.97	46	219356	59.97	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	119.94%
24) Chloroform-d	4.41	84	149518	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.60%
26) 1,2-Dichloroethane-d4	5.09	65	78665	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.60%
32) Benzene-d6	5.11	84	313398	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.40%
36) 1,2-Dichloropropane-d6	6.12	67	98857	4.84	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.80%
41) Toluene-d8	7.37	98	270089	4.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.40%
43) trans-1,3-Dichloropropene-	7.67	79	28392	3.11	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	62.20%
46) 2-Hexanone-d5	8.14	63	160103	40.61	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.22%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	62195	4.21	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	102927	4.77	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.40%

Target Compounds						Ovalue
33) Benzene	5.15	78	864381	11.10	ug/L	100
42) Toluene	7.44	91	198991	2.32	ug/L	100
51) Chlorobenzene	8.93	112	13966	0.25	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062218\
Data File : VV006418.D
Acq On : 22 Jun 2018 17:34
Operator : SY/MD
Sample : J3615-02 10X
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BEXT8DL

Quant Time: Jun 23 00:20:42 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 22 23:56:13 2018
Response via : Initial Calibration

