

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062218\
 Data File : VV006406.D
 Acq On : 22 Jun 2018 11:10
 Operator : SY/MD
 Sample : VV0622WBL01
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK61

Quant Time: Jun 23 00:00:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 22 23:56:13 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	350303	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	316768	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	141909	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	93550	4.89	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.80%
7) Chloroethane-d5	1.59	69	79979	5.22	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.40%
11) 1,1-Dichloroethene-d2	2.13	63	126556	3.40	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.00%
20) 2-Butanone-d5	3.97	46	203800	50.17	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.34%
24) Chloroform-d	4.41	84	162066	4.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.60%
26) 1,2-Dichloroethane-d4	5.09	65	82243	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.80%
32) Benzene-d6	5.10	84	357056	4.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.80%
36) 1,2-Dichloropropane-d6	6.12	67	106275	4.64	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.80%
41) Toluene-d8	7.36	98	322352	4.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.80%
43) trans-1,3-Dichloropropene-	7.67	79	34197	3.34	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	66.80%
46) 2-Hexanone-d5	8.14	63	154193	34.87	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	69.74%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	59067	3.56	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	71.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	114027	4.63	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062218\
Data File : VV006406.D
Acq On : 22 Jun 2018 11:10
Operator : SY/MD
Sample : VV0622WBL01
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK61

Quant Time: Jun 23 00:00:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 22 23:56:13 2018
Response via : Initial Calibration

