

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062218\
 Data File : VV006408.D
 Acq On : 22 Jun 2018 12:25
 Operator : SY/MD
 Sample : J3550-15
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Jun 23 00:03:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jun 22 23:56:13 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	305401	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	275122	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	120950	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	93860	5.63	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	112.60%
7) Chloroethane-d5	1.59	69	79978	5.99	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	119.80%
11) 1,1-Dichloroethene-d2	2.13	63	128470	3.96	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.20%
20) 2-Butanone-d5	3.97	46	209925	59.27	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.54%
24) Chloroform-d	4.41	84	164713	5.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.00%
26) 1,2-Dichloroethane-d4	5.09	65	83718	5.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.20%
32) Benzene-d6	5.11	84	361000	5.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.20%
36) 1,2-Dichloropropane-d6	6.12	67	108359	5.45	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.00%
41) Toluene-d8	7.37	98	322734	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.40%
43) trans-1,3-Dichloropropene-	7.67	79	33008	3.71	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.20%
46) 2-Hexanone-d5	8.14	63	156750	40.82	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.64%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	65262	4.53	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	115539	5.51	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062218\
Data File : VV006408.D
Acq On : 22 Jun 2018 12:25
Operator : SY/MD
Sample : J3550-15
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jun 23 00:03:16 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053118WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jun 22 23:56:13 2018
Response via : Initial Calibration

