

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV062718\
 Data File : VV006480.D
 Acq On : 27 Jun 2018 13:16
 Operator : SY/MD
 Sample : J3673-10
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DAZA5

Quant Time: Jun 28 01:24:52 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jun 28 01:18:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	319056	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	289838	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	127554	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	125861	6.00	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	120.00%
7) Chloroethane-d5	1.58	69	93126	5.51	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	110.20%
11) 1,1-Dichloroethene-d2	2.13	63	162427	4.37	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.40%
20) 2-Butanone-d5	3.96	46	211828	53.88	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.76%
24) Chloroform-d	4.41	84	182979	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.40%
26) 1,2-Dichloroethane-d4	5.09	65	89005	5.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.40%
32) Benzene-d6	5.11	84	413243	5.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.60%
36) 1,2-Dichloropropane-d6	6.12	67	119815	5.33	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.60%
41) Toluene-d8	7.37	98	373208	5.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.40%
43) trans-1,3-Dichloropropene-	7.67	79	35909	4.34	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.80%
46) 2-Hexanone-d5	8.14	63	155999	47.54	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.08%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	64741	4.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	126987	5.36	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.20%

Target Compounds

					Ovalue
5) Vinyl chloride	1.32	62	80409	3.09	ug/L 97
13) Acetone	2.21	43	9079	3.77	ug/L 99
18) trans-1,2-Dichloroethene	2.80	96	11366	0.54	ug/L 96
22) cis-1,2-Dichloroethene	3.97	96	1394936	60.05	ug/L 99
33) Benzene	5.15	78	26855	0.31	ug/L 100
34) Trichloroethene	5.96	95	292331	11.94	ug/L 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV062718\
Data File : VV006480.D
Acq On : 27 Jun 2018 13:16
Operator : SY/MD
Sample : J3673-10
Misc : 25 mL/MSVOA V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DAZA5

Quant Time: Jun 28 01:24:52 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR062518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jun 28 01:18:56 2018
Response via : Initial Calibration

