

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV070124\
 Data File : VV036395.D
 Acq On : 01 Jul 2024 23:52
 Operator : SY/MD
 Sample : P3087-14
 Misc : 5.59g/10.0mL/MSVOA_V/SOIL/A
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A4BQ8

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/05/2024
 Supervised By :Semsettin Yesilyurt 07/05/2024

Quant Time: Jul 02 05:26:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM062724SMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jul 02 05:19:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	322682	25.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	209260	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	43879	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.272	65	101244	18.475	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	73.880%		
7) Chloroethane-d5	1.526	69	103636	23.677	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	94.720%		
11) 1,1-Dichloroethene-d2	2.050	65	45612	17.705	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	70.840%		
21) 2-Butanone-d5	3.899	46	11738m	9.229	ug/L	0.06
Spiked Amount 50.000	Range 20 - 135		Recovery =	18.460%#		
24) Chloroform-d	4.243	84	206587	21.689	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	86.760%		
26) 1,2-Dichloroethane-d4	4.937	65	112025	22.054	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	88.200%		
32) Benzene-d6	4.953	84	343831	29.107	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	116.440%		
36) 1,2-Dichloropropane-d6	5.986	67	120466	33.904	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	135.600%#		
41) Toluene-d8	7.240	98	208596	19.585	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	78.360%		
43) trans-1,3-Dichloroprop...	7.561	79	16300	10.273	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	41.080%		
47) 2-Hexanone-d5	8.047	63	5851	10.316	ug/L	0.02
Spiked Amount 50.000	Range 20 - 135		Recovery =	20.640%		
56) 1,1,2,2-Tetrachloroeth...	10.153	84	69675	23.922	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	95.680%		
66) 1,2-Dichlorobenzene-d4	11.561	152	28552	17.856	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	71.440%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV070124\
Data File : VV036395.D
Acq On : 01 Jul 2024 23:52
Operator : SY/MD
Sample : P3087-14
Misc : 5.59g/10.0mL/MSVOA_V/SOIL/A
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A4BQ8

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/05/2024
Supervised By :Semsettin Yesilyurt 07/05/2024

Quant Time: Jul 02 05:26:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM062724SMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jul 02 05:19:46 2024
Response via : Initial Calibration

