

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071222\
 Data File : VV026718.D
 Acq On : 12 Jul 2022 11:41
 Operator : SY/MD
 Sample : VV0712WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK392

Quant Time: Jul 13 03:00:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR071122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jul 13 02:57:58 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	214298	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	217531	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	85347	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	104270	5.333	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 106.600%		
7) Chloroethane-d5	1.558	69	94042	5.567	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 111.400%		
11) 1,1-Dichloroethene-d2	2.098	63	151426	4.035	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 80.600%		
20) 2-Butanone-d5	3.925	46	165042	54.090	ug/L	0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	= 108.180%		
24) Chloroform-d	4.339	84	181527	5.310	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 106.200%		
26) 1,2-Dichloroethane-d4	5.031	65	84500	5.658	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 113.200%		
32) Benzene-d6	5.040	84	327390	4.953	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 99.000%		
36) 1,2-Dichloropropane-d6	6.069	67	106299	5.364	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 107.200%		
41) Toluene-d8	7.313	98	251952	4.294	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 85.800%		
43) trans-1,3-Dichloroprop...	7.622	79	30697	4.734	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 94.600%		
46) 2-Hexanone-d5	8.095	63	125402	44.669	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 89.340%		
56) 1,1,2,2-Tetrachloroeth...	10.214	84	76544	4.958	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 99.200%		
66) 1,2-Dichlorobenzene-d4	11.622	152	80140	5.389	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 107.800%		
Target Compounds						
16) Methylene chloride	2.500	84	4777	0.213	ug/L	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071222\
Data File : VV026718.D
Acq On : 12 Jul 2022 11:41
Operator : SY/MD
Sample : VV0712WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK392

Quant Time: Jul 13 03:00:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR071122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jul 13 02:57:58 2022
Response via : Initial Calibration

