

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW071222\
 Data File : VW026730.D
 Acq On : 12 Jul 2022 17:01
 Operator : SY/MD
 Sample : N3601-20DL 200X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBUB3DL

Quant Time: Jul 13 03:03:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR071122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jul 13 02:57:58 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	118443	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	119599	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	48219	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.298	65	96606	8.941	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 178.800%#		
7) Chloroethane-d5	1.561	69	85548	9.163	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 183.200%#		
11) 1,1-Dichloroethene-d2	2.098	63	137686	6.637	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 132.800%#		
20) 2-Butanone-d5	3.896	46	129368	76.711	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 153.420%#		
24) Chloroform-d	4.342	84	173819	9.200	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 184.000%#		
26) 1,2-Dichloroethane-d4	5.031	65	79106	9.584	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 191.600%#		
32) Benzene-d6	5.043	84	325588	8.959	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 179.200%#		
36) 1,2-Dichloropropane-d6	6.066	67	105877	9.717	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 194.400%#		
41) Toluene-d8	7.313	98	250754	7.773	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 155.400%#		
43) trans-1,3-Dichloroprop...	7.625	79	22523	6.318	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 126.400%		
46) 2-Hexanone-d5	8.088	63	137079	88.811	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 177.620%#		
56) 1,1,2,2-Tetrachloroeth...	10.214	84	76490	9.011	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 180.200%#		
66) 1,2-Dichlorobenzene-d4	11.622	152	80917	9.631	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 192.600%#		
Target Compounds						Qvalue
16) Methylene chloride	2.500	84	16603	1.342	ug/L	95
18) trans-1,2-Dichloroethene	2.751	96	2311	0.242	ug/L	82
22) cis-1,2-Dichloroethene	3.908	96	48340	5.021	ug/L	99
34) Trichloroethene	5.912	95	79445	8.084	ug/L	95
47) Tetrachloroethene	7.973	164	86160	11.179	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV071222\
Data File : VV026730.D
Acq On : 12 Jul 2022 17:01
Operator : SY/MD
Sample : N3601-20DL 200X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBUB3DL

Quant Time: Jul 13 03:03:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR071122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jul 13 02:57:58 2022
Response via : Initial Calibration

